

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007095

A. GENERAL Setup and Checks:

- Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
- Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
- Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
- Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-007095 DSP Ver: 1362.14
 - Core Ver: 8164.16 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 12763 s: 1 C: 47.0
9: 11441 s: 2 B: 44 0
(RFI indicator with keyed radio): 1
- Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
- Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 947 CF: 1.06
- Flow Monitor (Level 3,D,F): Flow Sensor, R: 199 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0
B: 44.9 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 638 psi P (gauge): 600 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: DR3377 Lot #: Milli QH₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: DR7111 Lot #: 17160 Exp date: 6-8-19
Known value: 0.020 AC
Results: 0.018 AC 0.019 AC 0.019 AC
3. High AC (Level 1,C):
Sim ser #: DR3375 Lot #: 17350 Exp date: 10-11-19
Known value: 0.300 AC
Results: 0.296 AC 0.297 AC 0.297 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR7089 Lot #: ICS#5 Exp date: 5.15.19
Known value: 0.10AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR3846 Lot #: ACE1 Exp date: NA
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3455 Lot #: MET2 Exp date: NA
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR5189 Lot #: IPA2 Exp date: NA
Known value: 0.10% ISOPROPRANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.

19. Stability Test (Level 3,T,S):

Ave: 0.0808 Std Dev: 0.0005 Rel Std Dev: 0.65%

Maintenance/Repairs/Notes: INITIAL INSPECTION AFTER
RECEIVING FROM CMI, INC.

Instrument is acceptable to be used in the field.

Charles E. Ehr

Field Inspector's Signature

10/15/18

Date

Roberta Muegen-Tumms

Reviewed by

22 Oct 18

Date

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 14:58

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuvw
ABCDEFGHIJKLMNOPQRSTUV

32 characters per line
abcdefghijklmnopqrstuvwxy
ABCDEFGHIJKLMNOPQRSTUVWXYZ

42 characters per line
abcdefghijklmnopqrstuvwxy
ABCDEFGHIJKLMNOPQRSTUVWXYZ


Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.081
Standard Lot #? 81
Standard Cyl #? 1111
Standard Expiration? 11/11/2019
Oper No? 666666

Flow Cal. Date: 09/27/2018
Slope 683
Intercept -637452

IR Calibration Date: 09/27/2018

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(x100): | -59629 | -23097 |
| 1st Coef(x100): | 252379 | 137307 |
| 2nd Coef(x100): | 2489 | 1549 |
| H2O adj(mg/l*10K): | 1414 | 594 |

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: *INTERNAL
PRINT TEST*

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 14:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.081
Standard Lot #? 81
Standard Cyl #? 1111
Standard Expiration? 11/11/2019
Oper No? 666666

Flow Cal. Date: 09/27/2018
Slope 683
Intercept -637452

IR Calibration Date: 09/27/2018
 3um 9um

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -59629 | -23097 |
| 1st Coef(*100): | 252379 | 137307 |
| 2nd Coef(*100): | 2489 | 1549 |
| H2O adj(mg/l*10k): | 1414 | 594 |

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

EXTERNAL
PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:01

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:02 |
| 02 *Subject Test | REF* | 15:05 |
| 03 Room Air | 0.000 | 15:05 |

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = R KEY REFUSAL
Dr. Lic. = ND/DIS821456
Lot No = 81
Cyl No = 1111
Expiration Date = 11/11/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: R KEY REFUSAL
BARCODE READER

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:05

DIAGNOSTIC

| | |
|---------------------------|------|
| Voltage/Current Test | Pass |
| RAM Test | Pass |
| EEPROM Checksum Test | Pass |
| Real Time Clock Test | Pass |
| DSP Test | Pass |
| Analytical Stability Test | Pass |
| Modem Test | Pass |
| Temp Regulation Test | Pass |

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:23

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 15:23 |
| 02 Std. Sol. | 0.000 | 15:24 |
| 03 Room Air | 0.000 | 15:25 |
| 04 Std. Sol. | 0.000 | 15:25 |
| 05 Room Air | 0.000 | 15:26 |
| 06 Std. Sol. | 0.000 | 15:27 |
| 07 Room Air | 0.000 | 15:27 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.000 AC
ZERO AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:35

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 15:36 |
| 02 Std. Sol. | 0.018 | 15:37 |
| 03 Room Air | 0.000 | 15:37 |
| 04 Std. Sol. | 0.019 | 15:38 |
| 05 Room Air | 0.000 | 15:38 |
| 06 Std. Sol. | 0.019 | 15:39 |
| 07 Room Air | 0.000 | 15:40 |

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17160

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:41

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 15:42 |
| 02 Std. Sol. | 0.296 | 15:43 |
| 03 Room Air | 0.000 | 15:43 |
| 04 Std. Sol. | 0.297 | 15:44 |
| 05 Room Air | 0.000 | 15:44 |
| 06 Std. Sol. | 0.297 | 15:45 |
| 07 Room Air | 0.000 | 15:46 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17350

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:48

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:49 |
| 02 *Subject Test | INT* | 15:50 |
| 03 Room Air | 0.000 | 15:50 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:53

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:54 |
| 02 *Subject Test | INT* | 15:54 |
| 03 Room Air | 0.000 | 15:55 |

*Invalid Test
Interferent Detected

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES851012
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:56

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:56 |
| 02 *Subject Test | INT* | 15:57 |
| 03 Room Air | 0.000 | 15:57 |

*Invalid Test
Interferent Detected

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/TES851012
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10% ISOPROPANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 15:58

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | AMB* | 15:59 |
| 02 Room Air | 0.000 | 16:00 |

*Invalid Test
Check Ambient Conditions

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/TES851012
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *AMBIENT FAIL*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:00

| Test | AC | Time |
|-------------------|-------|-------|
| 01 Room Air | 0.000 | 16:01 |
| 02 Invalid Sample | X.XXX | 16:01 |
| 03 Room Air | 0.000 | 16:02 |

*Invalid Test - Mouth Alcohol

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/TES851012
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INVALID SAMPLE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:02

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 16:03 |
| 02 *Subject Test | ABT* | 16:04 |
| 03 Room Air | 0.000 | 16:04 |

*Sequence Aborted

Sub Name = TEST, BOBBY NONE
Sub DOB = 01/01/1985
Sub Sex = Male Weight = NA
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/TES851012
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE
 ABORTED

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:09

| Test | AC | Time |
|-------------------|--------|-------|
| 01 Diagnostic | OK | 16:10 |
| 02 Room Air | 0.000 | 16:11 |
| 03 Subject Test 1 | 0.098 | 16:11 |
| 04 Room Air | 0.000 | 16:13 |
| 05 Std. Gas | 0.089* | 16:15 |
| 06 Room Air | 0.000 | 16:15 |

*Cal Check Out of Tolerance

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = CAL CHK OUT TOL
Dr. Lic. = ND/TES989643
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
OUT OF TOLERANCE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:18

| Test | AC | Time |
|------------------|--------|-------|
| 01 Diagnostic | OK | 16:19 |
| 02 Room Air | 0.000 | 16:19 |
| 03 *Subject Test | 0.000* | 16:22 |
| 04 Room Air | 0.000 | 16:24 |
| 05 Std. Gas | 0.081 | 16:25 |
| 06 Room Air | 0.000 | 16:27 |
| 07 *Subject Test | 0.000* | 16:30 |
| 08 Room Air | 0.000 | 16:31 |


No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/TES989643
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks: DEFICIENT
SAMPLES

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:32

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 16:32 |
| 02 Room Air | 0.000 | 16:33 |
| 03 Reference | IPS* | 16:33 |
| 04 Room Air | 0.000 | 16:33 |

*Improper Sample

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = IMPROPER SAMPLE
Dr. Lic. = ND/TES989643
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

*IMPROPER
SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:34

| Test | AC | Time |
|-------------------|--------|-------|
| 01 Diagnostic | OK | 16:34 |
| 02 Room Air | 0.000 | 16:35 |
| 03 Subject Test 1 | 0.251* | 16:35 |
| 04 Room Air | 0.000 | 16:38 |
| 05 Std. Gas | 0.082 | 16:39 |
| 06 Room Air | 0.000 | 16:40 |
| 07 Subject Test 2 | 0.097* | 16:41 |
| 08 Room Air | 0.000 | 16:41 |

*Difference Too Great

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/TES989643
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DIFFERENCE
TOO GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:42

| Test | AC | Time |
|-------------------|-------|-------|
| 01 Diagnostic | OK | 16:43 |
| 02 Room Air | 0.000 | 16:43 |
| 03 Subject Test 1 | 0.098 | 16:44 |
| 04 Room Air | 0.000 | 16:46 |
| 05 Std. Gas | 0.081 | 16:47 |
| 06 Room Air | 0.000 | 16:48 |
| 07 Subject Test 2 | 0.097 | 16:49 |
| 08 Room Air | 0.000 | 16:50 |
| 09 Reported AC | 0.097 | 16:49 |

Difference OK
No RFI Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = CMS MODE
Dr. Lic. = ND/TES989643
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *CMS MODE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
10/12/2018 16:51

| Test | AC | Time |
|---------------------|--------|-------|
| 01 Room Air | 0.000 | 16:51 |
| 02 Std. Gas | 0.081 | 16:51 |
| 03 Room Air | 0.000 | 16:52 |
| 04 Std. Gas | 0.080 | 16:52 |
| 05 Room Air | 0.000 | 16:53 |
| 06 Std. Gas | 0.081 | 16:53 |
| 07 Room Air | 0.000 | 16:54 |
| 08 Std. Gas | 0.081 | 16:54 |
| 09 Room Air | 0.000 | 16:55 |
| 10 Std. Gas | 0.080 | 16:55 |
| 11 Room Air | 0.000 | 16:55 |
| 12 Std. Gas | 0.081 | 16:56 |
| 13 Room Air | 0.000 | 16:56 |
| 14 Std. Gas | 0.082 | 16:57 |
| 15 Room Air | 0.000 | 16:57 |
| 16 Std. Gas | 0.080 | 16:58 |
| 17 Room Air | 0.000 | 16:58 |
| 18 Std. Gas | 0.081 | 16:58 |
| 19 Room Air | 0.000 | 16:59 |
| 20 Std. Gas | 0.081 | 16:59 |
| 21 Room Air | 0.000 | 17:00 |
| 22 Std. Gas | 0.081 | 17:00 |
| 23 Room Air | 0.000 | 17:01 |
| 24 Std. Gas | 0.081 | 17:01 |
| 25 Room Air | 0.000 | 17:02 |
| 26 Std. Gas | 0.081 | 17:02 |
| 27 Room Air | 0.000 | 17:02 |
| 28 Std. Gas | 0.080 | 17:03 |
| 29 Room Air | 0.000 | 17:03 |
| 30 Std. Gas | 0.080 | 17:04 |
| 31 Room Air | 0.000 | 17:04 |
| 32 Std. Gas | 0.081 | 17:04 |
| 33 Room Air | 0.000 | 17:05 |
| 34 Std. Gas | 0.081 | 17:05 |
| 35 Room Air | 0.000 | 17:06 |
| 36 Std. Gas | 0.081 | 17:06 |
| 37 Room Air | 0.000 | 17:07 |
| 38 Std. Gas | 0.081 | 17:07 |
| 39 Room Air | 0.000 | 17:08 |
| 40 Std. Gas | 0.081 | 17:08 |
| 41 Room Air | 0.000 | 17:08 |
| 42 Cal Check Stats: | | |
| 43 Average | 0.0808 | |
| 44 Std Dev | 0.0005 | |
| 45 Rel Std Dev | 0.65% | |



Operator Signature
CHARLES EDER

Remarks: STABILITY
TEST

Form 106-I8000

O.080 AC

LOT: 19817080A4

CYL: 021

EXP: 9/5/2019