

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-007093

Agency being assigned instrument: CMI, Inc.

Date 2/1/2019

Transferred from:

Name (print) CHARLES E. EDER

Signature Charles E. Eder

Agency ND O.A.G. - CRIMELAB

Transferred to:

Name (print) CMI, Inc.

Signature _____

Agency CMI, Inc

Reason for transfer/remarks:

- INSTRUMENT IS NOT FLAGGING MeOH @ 0.10% OR IS INTERMITTENTLY FLAGGING MeOH @ 0.10%
- BEING RETURNED TO CMI FOR EVALUATION

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____



Service Evaluation Form

This form **MUST** be completed and enclosed with instrument to be serviced.

Note: Please ship items in their original shipping container.

Contact information: Customer Number 585013 (contact Customer Service)
 Name CHARLES E. EDER Phone: (701) 328-6159
 Fax: (701) 328-6185 Email: ceder@nd.gov

Bill to Address:	Ship to Address:
<u>WARRANTY</u>	<u>ND O.A.G. - CRIME LAB</u>
	<u>ATTN: BREATH PROGRAM</u>
	<u>2641 E. MAIN AVE</u>
	<u>BISMARCK ND. 58501</u>

Instrument Serial Number: 80-007093

Detailed Description of Problem:
NOT FLAGGING MeOH INTERFERENT OR INTERMITTENTLY
FLAGGING MeOH @ 0.10% (SEE TEST RECORDS IN BOX)
 Hazardous Material Warning! - DO NOT return gas cylinders with instrument!

Note: An evaluation fee (\$83 (infrared) or \$43.⁵⁰ (fuel cell)) will apply to estimates which are not repaired.

I Authorize Repairs Up To: All \$250 \$500 \$750 Other \$ _____
 Purchase Order Number WARRANTY

After repair, add "Certificate of Calibration" for: S-D2, S-D5, I-300 (\$19)
 I-200/200D, I-240, I-400/400PA (\$39)

Authorized By:

CHARLES E. EDER STATE TOXICOLOGIST
 Name (Please Print) Title
Charles Eder 2/1/19
 Signature Date

Ship to:
 CMI, Inc.
 Attn: Service Dept.
 316 East Ninth Street
 Owensboro, KY 42303

No, please send estimate before repairs are made. An estimate will be faxed before performing any repairs and may cause delays in service.