

**CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT**

Instrument: 8000 Serial Number: 80-007092

Agency being assigned instrument: Crime Lab

Date 6 Feb 19

Transferred from:

Transferred to:

Name (print) Pam Tehl

Name (print) Roberta Grieger Nimmo

Signature [Signature]

Signature [Signature]

Agency Mercer City

Agency Crime Lab

Reason for transfer/remarks:

80-007092 is being exchanged for 80-006685.
80-007092 is being returned because the
Star + Test button doesn't always start the
countdown, but a re boot of the instrument will
get it into ready mode

Or Shipped Via:

USPS UPS FedEx Other

Date Sending Instrument: / / Air Bill No.