

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 11:58

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13518080A6
Standard Cyl #? 19
Standard Expiration? 08/05/2020
Oper No? 666666

Flow Cal. Date: 09/26/2018
Slope 672
Intercept -607823

IR Calibration Date: 09/26/2018
 3um 9um

0th Coef(*100): -28849 -15361
1st Coef(*100): 264571 134954
2nd Coef(*100): 3025 1343
H2O adj(mg/l*10k): 851 379

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:14
02 Std. Sol.	0.019	12:15
03 Room Air	0.000	12:15
04 Std. Sol.	0.019	12:16
05 Room Air	0.000	12:17
06 Std. Sol.	0.019	12:17
07 Room Air	0.000	12:18

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 19

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:24
02 Std. Sol.	0.297	12:25
03 Room Air	0.000	12:26
04 Std. Sol.	0.297	12:26
05 Room Air	0.000	12:27
06 Std. Sol.	0.298	12:28
07 Room Air	0.000	12:28


08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 19

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:04

Test	AC	Time
01 Room Air	0.000	12:04
02 *Subject Test	INT*	12:05
03 Room Air	0.000	12:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 19
Expiration Date = 08/05/2020
County = 19 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:06

Test	AC	Time
01 Diagnostic	OK	12:07
02 Room Air	0.000	12:07
03 *Subject Test	RFI*	12:08
04 Room Air	0.000	12:08

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 19
Expiration Date = 08/05/2020
County = 19 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:01
02 Std. Gas	0.080	12:01
03 Room Air	0.000	12:02
04 Std. Gas	0.080	12:02
05 Room Air	0.000	12:02
06 Std. Gas	0.079	12:03
07 Room Air	0.000	12:03

Lot No = 13518080A6
Cyl No = 19
Exp Date = 08/05/2020
County = 19 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:09
02 Std. Gas	0.080	12:09
03 Room Air	0.000	12:10
04 Std. Gas	0.079	12:10
05 Room Air	0.000	12:11
06 Std. Gas	0.079	12:11
07 Room Air	0.000	12:12

Lot No = 13518080A6
Cyl No = 19
Exp Date = 08/05/2020
County = 19 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
04/17/2019 12:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:31
02 Std. Gas	0.080	12:31
03 Room Air	0.000	12:32
04 Std. Gas	0.079	12:32
05 Room Air	0.000	12:33
06 Std. Gas	0.079	12:33
07 Room Air	0.000	12:34

Lot No = 13518080A6
Cyl No = 19
Exp Date = 08/05/2020
County = 19 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000