

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007097

Location: CASS COUNTY
CORRECTIONS
(CASS)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 608 psi Regulator: 600 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR5131 Lot #: 201810D Exp. Date: 10.24.20
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 13518080A6 Cyl #: 13 Exp. Date: 8.5.20

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

5.14.19
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:13

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13518080A6
Standard Cyl #? 13
Standard Expiration? 08/05/2020
Oper No? 666666

Flow Cal. Date: 09/27/2018
Slope 671
Intercept -612394

IR Calibration Date: 09/27/2018
 3um 9um

0th Coef(*100): -31035 -29197
1st Coef(*100): 248456 136978
2nd Coef(*100): 2388 1456
H2O adj(mg/l*10k): 964 610

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:26
02 Std. Sol.	0.018	19:27
03 Room Air	0.000	19:27
04 Std. Sol.	0.018	19:28
05 Room Air	0.000	19:28
06 Std. Sol.	0.018	19:29
07 Room Air	0.000	19:30

08 Sim Temp = 34.0°C

Simul Ser No = DR5131

Std Sol No = 201810D

County = 09

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:49
02 Std. Sol.	0.292	19:50
03 Room Air	0.000	19:51
04 Std. Sol.	0.293	19:51
05 Room Air	0.000	19:52
06 Std. Sol.	0.294	19:53
07 Room Air	0.000	19:53

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 09

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:11

Test	AC	Time
01 Room Air	0.000	19:12
02 *Subject Test	INT*	19:12
03 Room Air	0.000	19:13

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 13
Expiration Date = 08/05/2020
County = 09 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **INTERFERENT CHECK**
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:18

Test	AC	Time
01 Diagnostic	OK	19:19
02 Room Air	0.000	19:19
03 *Subject Test	RFI*	19:20
04 Room Air	0.000	19:20

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13518080A6
Cyl No = 13
Expiration Date = 08/05/2020
County = 09 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:14
02 Std. Gas	0.080	19:14
03 Room Air	0.000	19:15
04 Std. Gas	0.080	19:15
05 Room Air	0.000	19:16
06 Std. Gas	0.080	19:16
07 Room Air	0.000	19:17

Lot No = 13518080A6
Cyl No = 13
Exp Date = 08/05/2020
County = 09

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:21
02 Std. Gas	0.081	19:21
03 Room Air	0.000	19:22
04 Std. Gas	0.080	19:22
05 Room Air	0.000	19:23
06 Std. Gas	0.080	19:23
07 Room Air	0.000	19:24

Lot No = 13518080A6
Cyl No = 13
Exp Date = 08/05/2020
County = 09 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
05/14/2019 19:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:58
02 Std. Gas	0.081	19:58
03 Room Air	0.000	19:58
04 Std. Gas	0.080	19:59
05 Room Air	0.000	19:59
06 Std. Gas	0.080	20:00
07 Room Air	0.000	20:00

Lot No = 13518080A6
Cyl No = 13
Exp Date = 08/05/2020
County = 09 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **CALIBRATION CHECK**
0.080 AC

Form 106-I8000