



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Allen, Kathryn</i>		
Location <i>Cass Co. SO-Jail</i>	Intoxilyzer® Serial Number <i>80-007097</i>	
Gas Lot Number <i>13518080AC</i>	Gas Cylinder Number <i>13</i>	Gas Expiration Date <i>8-5-2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.081* AC  
 B. *0.080* AC  
 C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Kathryn Allen #3687</i>	Date <i>5-14-19</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Eche</i>	Date <i>5-20-2019</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007097  
Location = CASS      8164.16.00 09/18  
05/14/2019      10:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:41
02 Std. Gas	0.081	10:41
03 Room Air	0.000	10:42
04 Std. Gas	0.080	10:42
05 Room Air	0.000	10:42
06 Std. Gas	0.079	10:43
07 Room Air	0.000	10:43

Lot No = 13518080A6  
Cyl No = 13  
Exp Date = 08/05/2020  
County = 09      Oper No. = 090801

*Kathryn Allen #3687*  
Operator Signature  
KATHRYN ALLEN

Remarks:  
*New canister installed*

Form 106-I8000