



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Allen, Kathryn Marie</i>		
Location <i>Cass County Sheriffs Office - Jail</i>	Intoxilyzer® Serial Number <i>80-007097</i>	
Gas Lot Number <i>13518080A6</i>	Gas Cylinder Number <i>21</i>	Gas Expiration Date <i>8-5-2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.~~079~~ AC
 - B. 0.~~079~~ AC
 - C. 0.~~079~~ AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>K. Allen #3087</i>	Date <i>7-25-2019</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. [Signature]</i>	Date <i>7.25.19</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = CASS 8164.16.00 09/18
07/25/2019 01:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:48
02 Std. Gas	0.079	01:48
03 Room Air	0.000	01:48
04 Std. Gas	0.079	01:49
05 Room Air	0.000	01:49
06 Std. Gas	0.079	01:50
07 Room Air	0.000	01:50

Lot No = 13518080A6
Cyl No = 21
Exp Date = 08/05/2020
County = 09 Oper No. = 090801


Operator Signature
KATHRYN ALLEN

Remarks:
New Canister Install

Form 106-I8000

CASS COUNTY JAIL

Cass County Jail

450 34th St SW

Fargo, ND 58103

Booking : Tel: (701) 271-2918 Fax: (701) 271-2968

Date: 7-25-19

To: State Lab

Attention: C. Eder

Fax Number: 701-328-6185

From: Cass County Jail - Deputy K. Allen

Description: Gas Cylinder and ACA Test Reports for New gas canister install.

Number of pages (including cover): 3

If there are any problems receiving this transmission please call: 701-271-2918