



SERVICE WORK ORDER

DATE: 2/26/2020

316 E 9th Street / Owensboro KY 42303 / USA
Phone: 866-835-0690 Fax: 270-685-6268

404761

Bill To:
North Dakota, State Of
Crime Lab Division
2641 East Main Ave
Bismarck, ND 58501-5044
USA
Customer #: 585013
Fax: 701-328-2226

Ship To:
North Dakota, State Of
2641 East Main Ave
Crime Lab Division
Bismarck ND 58501-5044
USA

MODEL #: 002480ND

Serial Number: 80-007097

BILL CODE: Extended Warranty

EXTRA PARTS RCVD:

BT, BOX, BATTERY

DESCRIPTION OF PROBLEM

BREATH HOSE IS NOT HEATING

WORK PERFORMED:

breath hose/ erratic; o-ring/ cracked; CABLE ASSY,PRTR COM/ pinched; replaced parts listed; installed latest DSP software (02136216); calibrated/ final tested;

PARTS USED

Seq. No.	Part	Description	Quantity
60	471201	O-RING,BUNA-N,-008,3/16inx5/16	1.00 EA
80	650517	CERTIFICATE OF CALIBRATION	2.00 EA
90	441169	COVER DUST,5/8in x 1/2in	1.00 EA
100	470145	CAP,PLASTIC,.25IDx.50LG,RED	3.00 EA
110	470154	CAP,PLSTC,.406IDx.438-.562,RED	1.00 EA
120	470218	SEAL,TAMPER EVIDENT,SERVICE	1.00 EA
130	021307	HOSE ASSY,BREATH,I8000	1.00 EA
140	02136216	SOFTWARE, I8000 DSP, US MARKET	0.00 EA
150	330599	CABLE ASSY,PRTR COM,I8000	1.00 EA

LABOR / TESTING

Misc Code	Description	Hours
LABR	Service Repair Labor	
LCAL	Service - Calibration Adjust	0.50
LFT	Service - Cal / Final Test	1.00

Technician Name: Dewayne Varvel

SRV4047610001





316 E 9th Street / Owensboro KY 42303 / USA
Phone: 866-835-0690 Fax: 270-685-6268

SERVICE WORK ORDER

DATE: 2/26/2020

404761

Thank you for choosing CMI, Inc.

We wouldn't be here if it weren't for you...
And we know it.

Hopefully, our commitment to giving you a quality product was a key factor in your decision to purchase your Intoxilyzer. This is not where the quality ends; however, it is just where it begins. We pride ourselves on continuing to provide quality service and support through the life of this Intoxilyzer and all of your future Intoxilyzer purchases.

The ongoing quality service and support CMI provides can be accessed by dialing one number:

1-866-835-0690

If you have question about billing, shipping or other CMI products, Dialing 1-866-835-0690 can put you in touch with CMI Customer Service.

Dialing 1-866-835-0690 can also get answers to questions about setting up your instrument, installing software or servicing your instrument - just ask for one of our qualified service technicians.

CMI is an employee owned company and each one of us wants you to be a satisfied customer. We stand behind every one of our products. Our quality manufacturing, service and support is what sets us apart from the rest. Call us and you'll experience the difference.

Technician Name: Dewayne Varvel

SRV4047610001



SCTicket: SDH_PROD: Ver 1.7 (2020-01-16)
SN 80-007097

Page 2 of 3

Uploaded 5/28/2020 CEE

Page: 2 of 3



SERVICE WORK ORDER

DATE: 2/26/2020

316 E 9th Street / Owensboro KY 42303 / USA
Phone: 866-835-0690 Fax: 270-685-6268

404761

Statement of Warranty

Repair Product Warranty

Out of warranty product repairs are warranted for 90 days from the date of repair. This includes labor and those parts, which are replaced. If additional repair is required within the 90 day period, there will be a charge for any parts that were not originally replaced. Repairs made during the 90 day period that are unrelated to the original repair are not covered under the warranty.

There are no warranties expressed or implied, either oral or written, other than those contained in this warranty. In no event shall CMI be liable for any loss of profits or any indirect or consequential damages arising out of any such defect in material or workmanship.



Breath Alcohol Testing Is Our Business

316 East 9th Street
Owensboro, KY 42303
1-866-835-0690
www.alcoholtest.com

Technician Name: Dewayne Varvel

SRV4047610001



SCTicket: SDH-PROD:Ver 1.7 (2020-01-16)
SN 80-007097

Page 3 of 3

Uploaded 5/28/2020 CEE

Page: 3 of 3