

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007097

A. GENERAL Setup and Checks:

1. Power On
 5 min. countdown during initial warm-up
 1 min. countdown to begin test after standby mode
2. Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
 Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
4. Printer Setup (Level 3,S,P)
5. Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

1. Internal Printer Test and Paper Feed (Level 1,P). Attach record.
2. External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
 Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

1. Continuous Air Blank (Level 2,A)
2. Diagnostic (Level 2,D). Attach record.
3. Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
Inst Ser#: 80-007097 DSP Ver: 1362.16
Core Ver: 8164.16 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

1. DVM Monitor (Level 3,D,D): 3: 12381 s: 0 c: 47.00
9: 12938 s: 0 B: 45 0
(RFI indicator with keyed radio): 1
2. Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 962 CF: 1.05
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 196 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 47.5 I: 0.0
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 337 psi P (gauge): 350 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: MP3002 Lot #: MILLI Q H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: MP3064 Lot #: 20070 Exp date: 2.13.22
Known value: 0.020 AC
Results: 0.017 AC 0.017 AC 0.017 AC ^{CEE 0.017}
3. High AC (Level 1,C):
Sim ser #: MP3067 Lot #: 201911B Exp date: 11.5.21
Known value: 0.250 AC
Results: 0.248 AC 0.250 AC 0.250 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR3451 Lot #: ICS 6 Exp date: NA
Known value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR5188 Lot #: ACE 1 Exp date: NA
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3455 Lot #: MET 2 Exp date: NA
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR5189 Lot #: IPA 2 Exp date: NA
Known value: 0.10% ISOPROPANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0775 Std Dev: 0.0005 Rel Std Dev: 0.67%

Maintenance/Repairs/Notes: INITIAL INSPECTION DUE TO
FACTORY/WARRANTY REPAIR.

Instrument is acceptable to be used in the field.


Field Inspector's Signature

6.17.20
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:06

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMN

24 characters per line
abcdefghijklmnopqrstuvwx
ABCDEFGHIJKLMN

32 characters per line
abcdefghijklmnopqrstuvwxy123456
ABCDEFGHIJKLMN

42 characters per line
abcdefghijklmnopqrstuvwxy1234567890abcdef
ABCDEFGHIJKLMN

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.250
Standard Solution #? 2019118
Sim Serial #? MP3067
Oper No? 666666

Flow Cal. Date: 05/21/2020
Slope 689
Intercept -700113

IR Calibration Date: 05/21/2020
3um 9um

0th Coef(*100):	-20535	-26823
1st Coef(*100):	238050	136409
2nd Coef(*100):	4503	1573
H2O adj(mg/l)*10k:	571	497

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: *INTERNAL*
PRINT TEST

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:06

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.250
Standard Solution #? 201911B
Sim Serial #? MP3067
Oper No? 666666

Flow Cal. Date: 05/21/2020
Slope 689
Intercept -700113

IR Calibration Date: 05/21/2020

	3um	9um
0th Coef(*100):	-20535	-26823
1st Coef(*100):	238050	136409
2nd Coef(*100):	4503	1573
H2O adj(mg/l*10k):	571	497

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: EXTERNAL PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:09

Test	AC	Time
01 Room Air	0.000	16:09
02 *Subject Test	REF*	16:10
03 Room Air	0.000	16:10

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = R KEY REFUSAL
Dr. Lic. = ND/DIS921456
Simul Ser No = MP3067
Std Sol No = 201911B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: "R" KEY REFUSAL
BARCODE READER
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:10

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:14
02 Std. Sol.	0.000	16:14
03 Room Air	0.000	16:15
04 Std. Sol.	0.000	16:16
05 Room Air	0.000	16:16
06 Std. Sol.	0.000	16:17
07 Room Air	0.000	16:17

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: ZERO AC
0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Sol.	0.017	16:19
03 Room Air	0.000	16:20
04 Std. Sol.	0.017	16:21
05 Room Air	0.000	16:21
06 Std. Sol.	0.017	16:22
07 Room Air	0.000	16:22

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 20070

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/16/2020 16:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Sol.	0.248	16:25
03 Room Air	0.000	16:26
04 Std. Sol.	0.250	16:27
05 Room Air	0.000	16:27
06 Std. Sol.	0.250	16:28
07 Room Air	0.000	16:28

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = 201911B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.250 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 09:59

Test	AC	Time
01 Room Air	0.000	10:00
02 *Subject Test	INT*	10:00
03 Room Air	0.000	10:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:02

Test	AC	Time
01 Room Air	0.000	10:03
02 *Subject Test	INT*	10:03
03 Room Air	0.000	10:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:05

Test	AC	Time
01 Room Air	0.000	10:06
02 *Subject Test	INT*	10:06
03 Room Air	0.000	10:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:07

Test	AC	Time
01 Room Air	0.000	10:08
02 *Subject Test	INT*	10:09
03 Room Air	0.000	10:10

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10% ISOPROPANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:23

Test	AC	Time
01 Room Air	AMB*	10:24
02 Room Air	0.000	10:24

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

AMBIENT FAIL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:25

Test	AC	Time
01 Room Air	0.000	10:26
02 Invalid Sample	X.XXX	10:26
03 Room Air	0.000	10:27

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INVALID SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:28

Test	AC	Time
01 Room Air	0.000	10:28
02 *Subject Test	ABT*	10:29
03 Room Air	0.000	10:29

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE ABORTED

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:30

Test	AC	Time
01 Room Air	0.000	10:31
02 Subject Test 1	0.194	10:31
03 Room Air	0.000	10:32
04 Reported AC	0.194	10:31

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA MODE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:34

Test	AC	Time
01 Diagnostic	OK	10:35
02 Room Air	0.000	10:35
03 Subject Test 1	0.045	10:36
04 Room Air	0.000	10:38
05 Std. Gas	0.086*	10:39
06 Room Air	0.000	10:40

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL CHK OUT TOL
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
OUT OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:42

Test	AC	Time
01 Diagnostic	OK	10:42
02 Room Air	0.000	10:43
03 Subject Test 1	0.193	10:43
04 Room Air	0.000	10:46
05 Std. Gas	0.078	10:47
06 Room Air	0.000	10:48
07 *Subject Test	RFI*	10:48
08 Room Air	0.000	10:49

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 10:50

Test	AC	Time
01 Diagnostic	OK	10:51
02 Room Air	0.000	10:51
03 *Subject Test	0.000*	10:54
04 Room Air	0.000	10:56
05 Std. Gas	0.077	10:57
06 Room Air	0.000	10:59
07 *Subject Test	0.000*	11:02
08 Room Air	0.000	11:03

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks: DEFICIENT SAMPLES

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 11:04

Test	AC	Time
01 Diagnostic	OK	11:05
02 Room Air	0.000	11:06
03 Subject Test 1	0.045	11:06
04 Room Air	0.000	11:08
05 Std. Gas	0.078	11:09
06 Room Air	0.000	11:11
07 Reference	IPS*	11:11
08 Room Air	0.000	11:11

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = IMPROPER SAMP
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *IMPROPER SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 11:23

Test	AC	Time
01 Diagnostic	OK	11:24
02 Room Air	0.000	11:24
03 Subject Test 1	0.044*	11:24
04 Room Air	0.000	11:27
05 Std. Gas	0.078	11:28
06 Room Air	0.000	11:29
07 Subject Test 2	0.191*	11:30
08 Room Air	0.000	11:31

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DIFFERENCE TOO
GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 11:32

Test	AC	Time
01 Diagnostic	OK	11:33
02 Room Air	0.000	11:33
03 Subject Test 1	0.044	11:33
04 Room Air	0.000	11:36
05 Std. Gas	0.078	11:37
06 Room Air	0.000	11:38
07 Subject Test 2	0.044	11:39
08 Room Air	0.000	11:40
09 Reported AC	0.044	11:39

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS MODE
Dr. Lic. = ND/DIS921456
Lot No = 24119080A1
Cyl No = 9
Expiration Date = 11/05/2021
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

CMS MODE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = TOXL 8164.16.00 09/18
06/17/2020 11:41

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Gas	0.078	11:42
03 Room Air	0.000	11:42
04 Std. Gas	0.078	11:43
05 Room Air	0.000	11:43
06 Std. Gas	0.078	11:44
07 Room Air	0.000	11:44
08 Std. Gas	0.078	11:44
09 Room Air	0.000	11:45
10 Std. Gas	0.078	11:45
11 Room Air	0.000	11:46
12 Std. Gas	0.077	11:46
13 Room Air	0.000	11:47
14 Std. Gas	0.077	11:47
15 Room Air	0.000	11:48
16 Std. Gas	0.078	11:48
17 Room Air	0.000	11:49
18 Std. Gas	0.078	11:49
19 Room Air	0.000	11:49
20 Std. Gas	0.077	11:50
21 Room Air	0.000	11:50
22 Std. Gas	0.078	11:51
23 Room Air	0.000	11:51
24 Std. Gas	0.077	11:52
25 Room Air	0.000	11:52
26 Std. Gas	0.077	11:52
27 Room Air	0.000	11:53
28 Std. Gas	0.077	11:53
29 Room Air	0.000	11:54
30 Std. Gas	0.077	11:54
31 Room Air	0.000	11:55
32 Std. Gas	0.077	11:55
33 Room Air	0.000	11:56
34 Std. Gas	0.078	11:56
35 Room Air	0.000	11:56
36 Std. Gas	0.077	11:57
37 Room Air	0.000	11:57
38 Std. Gas	0.077	11:58
39 Room Air	0.000	11:58
40 Std. Gas	0.078	11:59
41 Room Air	0.000	11:59
42 Cal Check Stats:		
43 Average	0.0775	
44 Std Dev	0.0005	
45 Rel Std Dev	0.67%	



Operator Signature
CHARLES EDER

Remarks:

STABILITY TEST
0.080 AC

Form 106-I8000

LOT: 24119080 A1

CYL: 9

EXP: 11.5.21