


Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
06/30/2020 21:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:34
02 Std. Gas	0.078	21:34
03 Room Air	0.000	21:35
04 Std. Gas	0.078	21:35
05 Room Air	0.000	21:36
06 Std. Gas	0.078	21:36
07 Room Air	0.000	21:37

Lot No = 05620080A1
Cyl No = 37
Exp Date = 04/05/2022
County = 19 Oper No. = 132318


Operator Signature
TERRY DILLON

Remarks:

Form 106-18000