



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Terry Dillon</i>		
Location <i>Grant County S.D.</i>	Intoxilyzer® Serial Number <i>80-007094</i>	
Gas Lot Number <i>05620080 A1</i>	Gas Cylinder Number <i>37</i>	Gas Expiration Date <i>04/05/2022</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0. *078* AC
 B. 0. *078* AC
 C. 0. *078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Terry Dillon</i>	Date <i>06-30-2020</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ed</i>	Date <i>7.2.2020</i>


Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007094
Location = CARS 8164.16.00 09/18
06/30/2020 21:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:34
02 Std. Gas	0.078	21:34
03 Room Air	0.000	21:35
04 Std. Gas	0.078	21:35
05 Room Air	0.000	21:36
06 Std. Gas	0.078	21:36
07 Room Air	0.000	21:37

Lot No = 05620080A1
Cyl No = 37
Exp Date = 04/05/2022
County = 19 Oper No. = 132318


Operator Signature
TERRY DILLON

Remarks:

Form 106-18000