

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-007094

Agency being assigned instrument: O.A.G. - CRIME LAB - BREATH

Date 6/17/21

Transferred from:

Name (print) Terry Dillon

Signature [Signature]

Agency GRANT COUNTY S.O.

Transferred to:

Name (print) CHARLES E. EDER

Signature [Signature]

Agency O.A.G. - CRIME LAB - BREATH

Reason for transfer/remarks:

- TRANSPORTED TO CRIME LAB FOR ANNUAL INSPECTION
- NOTE: 80-007094 EXHIBITED EEPROM ERROR, EEPROM FAULT, SYSTEM FAULT ID #018 CEE TD
- BEING REPLACED BY 80-004200

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____