



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-007097	Instrument Location Rolette County S.O
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Annual Inspection
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 4-21-22
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By 	Certified Date 11 May 2022
---------------------------	--------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = ROLL 8164.16.00 09/18
04/12/2022 12:21

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34921080A1
Standard Cyl #? 61
Standard Expiration? 02/05/2024
Oper No? 130918

Flow Cal. Date: 09/03/2020
Slope 660
Intercept -636264

IR Calibration Date: 09/03/2020
 3um 9um

0th Coef(*100): -36810 -25919
1st Coef(*100): 254044 139068
2nd Coef(*100): 1858 1258
H2O adj(mg/l*10k): 903 584

***** Printer Test End *****



Operator Signature
MITCH SLATER

Remarks: **Print Test**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = ROLL 8164.16.00 09/18
04/12/2022 12:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:47
02 Std. Gas	0.082	12:48
03 Room Air	0.000	12:48
04 Std. Gas	0.083	12:48
05 Room Air	0.000	12:49
06 Std. Gas	0.083	12:49
07 Room Air	0.000	12:50

Lot No = 34921080A1
Cyl No = 61
Exp Date = 02/05/2024
County = 40 Oper No. = 130918


Operator Signature
MITCH SLATER

Remarks: ACA TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = ROLL 8164.16.00 09/18
04/12/2022 12:51

Test	AC	Time
01 Room Air	0.000	12:54
02 *Subject Test	RFI*	12:54
03 Room Air	0.000	12:54

*Invalid Test
 Inhibited - RFI

Sub Name = RFI TEST, INSTALL NA
Sub DOB = 04/12/1900
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 34921080A1
Cyl No = 61
Expiration Date = 02/05/2024
County = 40 Oper No. = 130918



Operator Signature
MITCH SLATER

Remarks: RFI TEST