

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-067092

Location: TUXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 863 psi Regulator: 850 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP6036 Lot #: 21340 Exp. Date: 8/24/2023
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3068 Lot #: 202201F Exp. Date: 1/18/2024

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: ICSC Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual inspection.

Replaced breath tube o-ring. Replaced external battery.
Audible tone on the instrument is quite but still can be heard.

Instrument is acceptable to be used in the field.

Jemelle Pertschiller
Breath Analyst Signature

25 May 22
Date

[Signature]
Reviewed by

052622
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 11:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 17
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 09/26/2018
Slope 670
Intercept -596123

IR Calibration Date: 09/26/2018
 3um 9um

0th Coef(*100): -27453 -17291
1st Coef(*100): 248629 134601
2nd Coef(*100): 2770 1393
H2O adj(mg/l*10k): 872 414

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 11:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Sol.	0.017	11:43
03 Room Air	0.000	11:44
04 Std. Sol.	0.017	11:44
05 Room Air	0.000	11:45
06 Std. Sol.	0.017	11:46
07 Room Air	0.000	11:46

08 Sim Temp = 34.0°C

Simul Ser No = MP6036
Std Sol No = 21340
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 11:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:48
02 Std. Sol.	0.290	11:49
03 Room Air	0.000	11:50
04 Std. Sol.	0.292	11:50
05 Room Air	0.000	11:51
06 Std. Sol.	0.292	11:52
07 Room Air	0.000	11:52

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 202201F
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *High AC 0.300*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 11:55

Test	AC	Time
01 Room Air	0.000	11:56
02 *Subject Test	INT*	11:56
03 Room Air	0.000	11:57

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portschieller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 11:57

Test	AC	Time
01 Diagnostic	OK	11:58
02 Room Air	0.000	11:58
03 *Subject Test	RFI*	12:00
04 Room Air	0.000	12:00

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 12:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:02
02 Std. Gas	0.080	12:02
03 Room Air	0.000	12:03
04 Std. Gas	0.080	12:03
05 Room Air	0.000	12:04
06 Std. Gas	0.080	12:04
07 Room Air	0.000	12:04

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 12:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Gas	0.080	12:06
03 Room Air	0.000	12:06
04 Std. Gas	0.080	12:06
05 Room Air	0.000	12:07
06 Std. Gas	0.080	12:07
07 Room Air	0.000	12:08

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007092
Location = TOXL 8164.16.00 09/18
05/25/2022 12:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:09
02 Std. Gas	0.080	12:09
03 Room Air	0.000	12:10
04 Std. Gas	0.080	12:10
05 Room Air	0.000	12:11
06 Std. Gas	0.080	12:11
07 Room Air	0.000	12:12

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration checks^{for} check # 7-9*