



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-007096	Instrument Location BE PD
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Yearly Certification @ Crime Lab
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>ALM 1745</i>	Date <i>6-17-22</i>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Pertschella</i>	Certified Date <i>06/23/22</i>
---	-----------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007096
Location = BEPD 8164.16.00 09/18
06/17/2022 14:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 26021080A1
Standard Cyl #? 36
Standard Expiration? 10/05/2023
Oper No? 131958

Flow Cal. Date: 10/02/2018
Slope 695
Intercept -692605

IR Calibration Date: 10/02/2018
 3um 9um

0th Coef(*100): -33544 -18475
1st Coef(*100): 261464 138054
2nd Coef(*100): 3020 1612
H2O adj(mg/l*10k): 573 386

***** Printer Test End *****

Baker 1745

Operator Signature
ADAM BAKER

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007096
Location = BEPD 8164.16.00 09/18
06/17/2022 14:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:29
02 Std. Gas	0.079	14:29
03 Room Air	0.000	14:30
04 Std. Gas	0.080	14:30
05 Room Air	0.000	14:30
06 Std. Gas	0.079	14:31
07 Room Air	0.000	14:31

Lot No = 26021080A1
Cyl No = 36
Exp Date = 10/05/2023
County = 29 Oper No. = 131958



Operator Signature
ADAM BAKER

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007096
Location = BEPD 8164.16.00 09/18
06/17/2022 14:35

Test	AC	Time
01 Room Air	RFI*	14:36
02 Room Air	RFI*	14:36

*Invalid Test
Inhibited - RFI

Sub Name = BAKER, ADAM JAMES
Sub DOB = 06/29/1982
Sub Sex = Male Weight = 190
Test = MIC Cit = 0000000000000000
Dr. Lic. = ND/BAK823805
Lot No = 26021080A1
Cyl No = 36
Expiration Date = 10/05/2023
County = 29 Oper No. = 131958



Operator Signature
ADAM BAKER

Remarks: