

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007091 Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 1005 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: 1656 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.078 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1	<u>0.078</u>	Test 4	<u>0.078</u>	Test 7	<u>0.078</u>
Test 2	<u>0.078</u>	Test 5	<u>0.078</u>	Test 8	<u>0.078</u>
Test 3	<u>0.078</u>	Test 6	<u>0.078</u>	Test 9	<u>0.078</u>
Average	<u>0.078</u>				

C. Remarks/Maintenance: Annual inspection.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Portschiller
Breath Analyst Signature

6/23/22
Date

[Signature]
Reviewed by

062422
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
06/23/2022 12:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:55
02 Std. Sol.	0.018	12:56
03 Room Air	0.000	12:57
04 Std. Sol.	0.018	12:57
05 Room Air	0.000	12:58
06 Std. Sol.	0.018	12:59
07 Room Air	0.000	12:59


08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks:

Low AC check 0.020

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
06/23/2022 13:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:06
02 Std. Sol.	0.295	13:07
03 Room Air	0.000	13:08
04 Std. Sol.	0.296	13:09
05 Room Air	0.000	13:09
06 Std. Sol.	0.297	13:10
07 Room Air	0.000	13:10

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks:

High Ac Check 0.300


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
06/23/2022 13:12

Test	AC	Time
01 Room Air	0.000	13:13
02 *Subject Test	INT*	13:14
03 Room Air	0.000	13:15

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/TES989643
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
06/23/2022 13:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Gas	0.078	13:19
03 Room Air	0.000	13:20
04 Std. Gas	0.078	13:20
05 Room Air	0.000	13:21
06 Std. Gas	0.078	13:21
07 Room Air	0.000	13:22

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check 41-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
06/23/2022 13:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:22
02 Std. Gas	0.078	13:23
03 Room Air	0.000	13:23
04 Std. Gas	0.078	13:24
05 Room Air	0.000	13:24
06 Std. Gas	0.078	13:25
07 Room Air	0.000	13:25

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007091
Location = TOXL 8164.16.00 09/18
06/23/2022 13:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:26
02 Std. Gas	0.078	13:26
03 Room Air	0.000	13:27
04 Std. Gas	0.078	13:27
05 Room Air	0.000	13:28
06 Std. Gas	0.078	13:28
07 Room Air	0.000	13:28

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: Dry Calibration Check #1-3 ^{#7-9} JLP