

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-0070915 Location: TOXL

A. General Setup and Checks:

1. ☒ Diagnostic Tests Pass and Instrument in Ready Mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 994 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

☒ Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. ☒ Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP5290 Lot #: 21340 Exp. Date: 8/24/23
- b. ☒ High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3067 Lot #: 21420 Exp. Date: 10/20/23

☒ Configure dry gas standard for the following tests (Level 1,S).

2. ☒ Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7345 Lot #: 1LS7 Exp. Date: NA
☒ Display reads "Interferent Detect"
3. ☒ RFI Check (CMS Mode) ☒ Display reads "RFI Detect"
4. ☒ Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 31 Exp. Date: 3/5/23

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual inspection.

☒ Instrument is acceptable to be used in the field.

Janelle Portchiller
Breath Analyst Signature

06/30/22
Date

[Signature]
Reviewed by

070122
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 12:48

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 3
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 09/27/2018
Slope 683
Intercept -637452

IR Calibration Date: 09/27/2018
 3um 9um

0th Coef(*100):	-59629	-23097
1st Coef(*100):	252379	137307
2nd Coef(*100):	2489	1549
H2O adj (mg/l*10k):	1414	594

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 12:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Sol.	0.018	12:59
03 Room Air	0.000	13:00
04 Std. Sol.	0.018	13:01
05 Room Air	0.000	13:01
06 Std. Sol.	0.018	13:02
07 Room Air	0.000	13:02

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 21340

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC check 0.020*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 13:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Sol.	0.297	13:04
03 Room Air	0.000	13:05
04 Std. Sol.	0.299	13:06
05 Room Air	0.000	13:06
06 Std. Sol.	0.299	13:07
07 Room Air	0.000	13:07

08 Sim Temp = 34.0°C

Simul Ser No = MP3067

Std Sol No = 21420

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 13:10

Test	AC	Time
01 Room Air	0.000	13:11
02 *Subject Test	INT*	13:12
03 Room Air	0.000	13:12

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 13:13

Test	AC	Time
01 Diagnostic	OK	13:13
02 Room Air	0.000	13:14
03 *Subject Test	RFI*	13:14
04 Room Air	0.000	13:15

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 31
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 13:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:16
02 Std. Gas	0.081	13:17
03 Room Air	0.000	13:17
04 Std. Gas	0.081	13:17
05 Room Air	0.000	13:18
06 Std. Gas	0.081	13:18
07 Room Air	0.000	13:19


Lot No = 02621080A1

Cyl No = 31

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 13:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:20
02 Std. Gas	0.081	13:20
03 Room Air	0.000	13:21
04 Std. Gas	0.081	13:21
05 Room Air	0.000	13:22
06 Std. Gas	0.082	13:22
07 Room Air	0.000	13:22

Lot No = 02621080A1
Cyl No = 31
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #4-6*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007095
Location = TOXL 8164.16.00 09/18
06/30/2022 13:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:24
02 Std. Gas	0.081	13:25
03 Room Air	0.000	13:25
04 Std. Gas	0.082	13:25
05 Room Air	0.000	13:26
06 Std. Gas	0.082	13:26
07 Room Air	0.000	13:27

Lot No = 02621080A1

Cyl No = 31

Exp Date = 03/05/2023

County = 08

Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*