



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-007093	Instrument Location BOPD
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000. *Fax or N/A*
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): *Annual inspection*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>J. By...</i>	Date 6.30.2022
--	--------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Janelle Portschiller</i>	Certified Date 07/01/2022
--	-------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007093
Location = BOPD 8164.16.00 09/18
06/30/2022 16:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:28
02 Std. Gas	0.078	16:28
03 Room Air	0.000	16:29
04 Std. Gas	0.078	16:29
05 Room Air	0.000	16:30
06 Std. Gas	0.078	16:30
07 Room Air	0.000	16:31

Lot No = 02621080A1
Cyl No = 24
Exp Date = 03/05/2023
County = 06 Oper No. = 130538

T. Bergerson #1821
Operator Signature
TREVOR BERGERSON

Remarks:
Install.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007093
Location = BOPD 8164.16.00 09/18
06/30/2022 16:32

Test	AC	Time
01 Diagnostic	OK	16:34
02 Room Air	RFI*	16:34
03 Room Air	0.000	16:34

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST RFI TEST
Sub DOB = 06/03/1988
Sub Sex = Male Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 02621080A1
Cyl No = 24
Expiration Date = 03/05/2023
County = 06 Oper No. = 130538

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

T. Bergerson #1821

Operator Signature
TREVOR BERGERSON

Remarks:

Install.