

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006510

1) GENERAL Setup and Checks:

- a) Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
- b) Display Intensity (Ctrl + ↑ or ↓)
- c) Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
- d) Printer Setup (Level 3,S,P)
- e) Forms Load (CMI, Inc., CoBrA Software)

2) LEVEL 1 Setup and Checks:

- a) Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- b) External Printer Test (Level 1,P). Attach record.
- c) Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

3) LEVEL 2 Setup and Checks:

- a) Continuous Air Blank (Level 2,A)
- b) Diagnostic (Level 2,D). Attach record.
- c) Prelim Data Entry (Level 2,E)
- d) Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-006510 DSP Ver: 1362.14
 - Core Ver: 8164.13 CPLD Ver: 7

4) LEVEL 3 Setup and Checks:

- a) DVM Monitor (Level 3,D,D): 3: 12473 s: 2 C: 47.00
9: 13231 s: 0 B: 45 0
(RFI indicator with keyed radio): 1
- b) Function (F) Keys: : F5 – Tone F9 – Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
- c) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 957 CF: 1.05
- d) Flow Monitor (Level 3,D,F): Flow Sensor, R: 175 F: 0,000 L/s

- e) Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.5 I: 0.0
B: 45.0 V: 0.0
- f) Tank Monitor (Level 3,D,G): P (display): 734 psi P (gauge): ≈740 psi
- g) General Setup (Level 3,S,G)
- h) Configure Start Test (Level 3,S,T)
- i) Network (Level 3,S,N)

5) TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

- a) Zero AC (Level 1,C):
Sim ser #: DR3375 Lot #: N/A (Milli Q) Exp date: N/A
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
- b) Low AC (Level 1,C):
Sim ser #: DR3344 Lot #: 14240 Exp date: 10/20/16
Known value: 0.020 AC
Results: 0.019 AC 0.019 AC 0.019 AC
- c) High AC (Level 1,C):
Sim ser #: DR3841 Lot #: 201408C Exp date: 8/12/16
Known value: 0.300 AC
Results: 0.292 AC 0.292 AC 0.292 AC

Configure dry gas standard before running tests (Level 1,S):

- d) Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR5144 Lot #: 1cs4 Exp date: 09/09/2017
Known value: 0.10 AC + 0.05% acetone
 Display reads "Interferent Detect"
- e) Acetone (Level 1,B):
Sim ser #: DR3455 Lot #: Acc 1 Exp date: 09/29/2017
Known value: 0.05% acetone
 Display reads "Interferent Detect"
- f) Methanol (Level 1,B):
Sim ser #: DR5189 Lot #: Met 2 Exp date: 09/29/2017
Known value: 0.10% methanol
 Display reads "Interferent Detect"
- g) Isopropanol (Level 1,B):
Sim ser #: DR3454 Lot #: Ipa 2 Exp date: 09/29/2017
Known value: 0.10% isopropanol
 Display reads "Interferent Detect"

- h) Ambient Fail (Level 1,B): Display and test record read correct.
- i) Invalid Sample (Level 1,B): Display and test record read correct.
- j) Sequence Aborted (Level 1,B): Display and test record read correct.
- k) ABA (Level 1,B): Test record reads correct.
- l) Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- m) RFI Detected (B,C, or CMS mode): Display and test record read correct.
- n) Tank Pressure Below Minimum (CMS mode): Display reads correct.
- o) Deficient Samples (CMS mode): Display and test record read correct.
- p) Improper Sample (CMS mode): Display and test record read correct.
- q) Difference Too Great (CMS mode): Display and test record read correct.
- r) CMS (CMS mode): Test record reads correct.
- s) Stability Test (Level 3,T,S):

Ave: 0.0805 Std Dev: 0.0005 Rel Std Dev: 0.64%

Maintenance/Repairs/Notes: _____

Instrument is acceptable to be used in the field.

Deb Shanaver
Field Inspector's Signature

20 Jan 2016
Date

[Signature]
Reviewed by

1/20/16
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/14/2016 15:05

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuvw
ABCDEFGHIJKLMNQRSTUW

32 characters per line
abcdefghijklmnopqrstuvwxy123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuvwxy1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menful? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 80
Standard Cyl #? 13478
Standard Expiration? 11/11/2020
Oper No? 777777

Flow Cal. Date: 08/24/2015
Slope 624
Intercept -539297

IR Calibration Date: 08/24/2015

	3um	9um
0th Coef(*100):	-23613	-13697
1st Coef(*100):	251959	131614
2nd Coef(*100):	4659	1853
H2O adj(mg/l*10K):	819	493

***** Test End *****

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Print test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 14:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
|
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
|
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 25
Standard Expiration? 12/05/2016
Oper No? 777777

Flow Cal. Date: 08/24/2015
Slope 624
Intercept -539297

IR Calibration Date: 08/24/2015
3um 9um

	3um	9um
0th Coef(*100):	-23613	-13697
1st Coef(*100):	251959	131614
2nd Coef(*100):	4659	1853
H2O adj(mg/l*10k):	819	493

***** Printer Test End *****


Operator Signature
DEB SHANAVER

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 14:05

Test	AC	Time
01 Room Air	0.000	14:06
02 *Subject Test	REF*	14:06
03 Room Air	0.000	14:07

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Refusal

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 14:08

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Diagnostic test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 14:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Sol.	0.000	14:59
03 Room Air	0.000	14:59
04 Std. Sol.	0.000	15:00
05 Room Air	0.000	15:00
06 Std. Sol.	0.000	15:01
07 Room Air	0.000	15:01

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = MILLIQ

County = 08

Oper No. = 77777



Operator Signature

DEB SHANAVER

Remarks:

Zero AC 0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 14:19

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:19
02 Std. Sol.	0.019	14:20
03 Room Air	0.000	14:21
04 Std. Sol.	0.019	14:22
05 Room Air	0.000	14:22
06 Std. Sol.	0.019	14:23
07 Room Air	0.000	14:24

08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 14240

County = 08

Oper No. = 77777

Deb Shauver
Operator Signature
DEB SHANAVER

Remarks:

low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 14:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:33
02 Std. Sol.	0.292	14:33
03 Room Air	0.000	14:34
04 Std. Sol.	0.292	14:35
05 Room Air	0.000	14:36
06 Std. Sol.	0.292	14:36
07 Room Air	0.000	14:37

08 Sim Temp = 34.0°C

Simul Ser No = DR3841
Std Sol No = 201408C
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

High AC 0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 16:03

Test	AC	Time
01 Room Air	0.000	16:04
02 *Subject Test	INT*	16:04
03 Room Air	0.000	16:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10 AC + 0.05 g acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 16:12

Test	AC	Time
01 Room Air	0.000	16:13
02 *Subject Test	INT*	16:13
03 Room Air	0.000	16:14

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 16:26

Test	AC	Time
01 Room Air	0.000	16:27
02 *Subject Test	INT*	16:27
03 Room Air	0.000	16:28

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.1090 Methanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 11:16

Test	AC	Time
01 Room Air	0.000	11:17
02 *Subject Test	INT*	11:17
03 Room Air	0.000	11:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.1090 isopropanal

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 12:52

Test	AC	Time
01 Room Air	AMB*	12:53
02 Room Air	0.000	12:54

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Ambient fail

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 13:08

Test	AC	Time
01 Room Air	0.000	13:09
02 Invalid Sample	X.XXX	13:10
03 Room Air	0.000	13:10

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Invalid sample - mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 13:11

Test	AC	Time
01 Room Air	0.000	13:12
02 *Subject Test	ABT*	13:12
03 Room Air	0.000	13:13

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Sequence aborted

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 13:18

Test	AC	Time
01 Room Air	0.000	13:18
02 Subject Test 1	0.007	13:19
03 Room Air	0.000	13:19
04 Reported AC	0.007	13:19

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

ABA test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 13:41

Test	AC	Time
01 Diagnostic	OK	13:41
02 Room Air	0.000	13:42
03 Subject Test 1	0.000	13:42
04 Room Air	0.000	13:44
05 Std. Gas	0.100*	13:45
06 Room Air	0.000	13:46

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

Calibration check out of tolerance

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 13:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	13:57
02 Room Air	0.000	13:58

*Invalid Test
Inhibited - RFI

Lot No = 29614080A2
Cyl No = 25
Exp Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 13:59

Test	AC	Time
01 Diagnostic	OK	14:00
02 Room Air	0.000	14:00
03 *Subject Test	0.000*	14:03
04 Room Air	0.000	14:05
05 Std. Gas	0.082	14:07
06 Room Air	0.000	14:08
07 *Subject Test	0.000*	14:11
08 Room Air	0.000	14:12

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777


20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 14:14

Test	AC	Time
01 Diagnostic	OK	14:15
02 Room Air	0.000	14:15
03 *Subject Test	0.000*	14:18
04 Room Air	IPS**	14:19
05 Room Air	0.000	14:20

*Deficient Sample - Value Printed was
Highest Obtained
**Improper Sample


Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 14:20

Test	AC	Time
01 Diagnostic	OK	14:21
02 Room Air	0.000	14:21
03 Subject Test 1	0.042*	14:22
04 Room Air	0.000	14:24
05 Std. Gas	0.082	14:25
06 Room Air	0.000	14:27
07 Subject Test 2	0.000*	14:27
08 Room Air	0.000	14:28

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
DEB SHANAVER

Remarks:

Difference too great

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/20/2016 14:40

Test	AC	Time
01 Diagnostic	OK	14:41
02 Room Air	0.000	14:42
03 Subject Test 1	0.000	14:42
04 Room Air	0.000	14:44
05 Std. Gas	0.081	14:45
06 Room Air	0.000	14:46
07 Subject Test 2	0.000	14:47
08 Room Air	0.000	14:47
09 Reported AC	0.000	14:42

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 29614080A2
Cyl No = 25
Expiration Date = 12/05/2016
County = 08 Oper No. = 777777

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

CMS

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006510
Location = TOXL 8164.13.00 06/09
01/19/2016 13:32

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Gas	0.080	13:33
03 Room Air	0.000	13:34
04 Std. Gas	0.080	13:34
05 Room Air	0.000	13:35
06 Std. Gas	0.080	13:35
07 Room Air	0.000	13:36
08 Std. Gas	0.081	13:36
09 Room Air	0.000	13:37
10 Std. Gas	0.081	13:38
11 Room Air	0.000	13:38
12 Std. Gas	0.080	13:39
13 Room Air	0.000	13:39
14 Std. Gas	0.080	13:40
15 Room Air	0.000	13:40
16 Std. Gas	0.080	13:41
17 Room Air	0.000	13:41
18 Std. Gas	0.080	13:42
19 Room Air	0.000	13:42
20 Std. Gas	0.080	13:43
21 Room Air	0.000	13:43
22 Std. Gas	0.081	13:44
23 Room Air	0.000	13:45
24 Std. Gas	0.081	13:45
25 Room Air	0.000	13:46
26 Std. Gas	0.081	13:46
27 Room Air	0.000	13:47
28 Std. Gas	0.081	13:47
29 Room Air	0.000	13:48
30 Std. Gas	0.080	13:48
31 Room Air	0.000	13:49
32 Std. Gas	0.081	13:49
33 Room Air	0.000	13:50
34 Std. Gas	0.081	13:50
35 Room Air	0.000	13:51
36 Std. Gas	0.081	13:51
37 Room Air	0.000	13:52
38 Std. Gas	0.080	13:52
39 Room Air	0.000	13:53
40 Std. Gas	0.080	13:53
41 Room Air	0.000	13:54
42 Cal Check Stats:		
43 Average	0.0805	
44 Std Dev	0.0005	
45 Rel Std Dev	0.64%	

*Deb Shanaver
Stability Test*

Operator Signature
DEB SHANAVER

Remarks:

lot: 2961480 A2

cyt: 25

exp: 12/05/2016

Form 106-I8000

Deb Shanaver
Stability test