

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006512

A. GENERAL Setup and Checks:

1. Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
2. Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
4. Printer Setup (Level 3,S,P)
5. Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

1. Internal Printer Test and Paper Feed (Level 1,P). Attach record.
2. External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

1. Continuous Air Blank (Level 2,A)
2. Diagnostic (Level 2,D). Attach record.
3. Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-006512 DSP Ver: 1362.14
 - Core Ver: 8164.13 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

1. DVM Monitor (Level 3,D,D): 3: 12734 s: 1 c: 47.00
9: 13596 s: 1 B: 45 0
(RFI indicator with keyed radio): 1
2. Function (F) Keys : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 955 CF: 1.05
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 231 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 1002 psi P (gauge): ~1000 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: MP3067 Lot #: MILLI Q H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: MP3066 Lot #: 15160 Exp date: 7/14/17
Known value: 0.020 AC
Results: 0.020 AC 0.020 AC 0.020 AC
3. High AC (Level 1,C):
Sim ser #: MP3068 Lot #: 14102 Exp date: 4/23/16
Known value: 0.400 AC CEE
Results: 0.394 AC 0.396⁴ AC 0.396 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR5135 Lot #: ICS4 Exp date: 9/9/2017
Known value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR5143 Lot #: ACE2 Exp date: 9/29/17
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3377 Lot #: MET2 Exp date: 9/29/17
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR3453 Lot #: IPA2 Exp date: 9/29/17
Known value: 0.10% ISOPROPRANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.

19. Stability Test (Level 3,T,S):

Ave: 0.0815 Std Dev: 0.0005 Rel Std Dev: 0.63%

Maintenance/Repairs/Notes: _____

Instrument is acceptable to be used in the field.

Charles E. Edr
Field Inspector's Signature

21 JAN 16
Date

Deb Shanauer
Reviewed by

21 Jan 16
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/20/2016 16:33

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNOPQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNOPQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNOPQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 6
Standard Expiration? 12/05/2016
Oper No? 666666

Flow Cal. Date: 08/20/2015
Slope: 680
Intercept: -660856

IR Calibration Date: 08/21/2015

	3um	9um
0th Coef(*100):	-23051	-19261
1st Coef(*100):	266836	137135
2nd Coef(*100):	3974	1466
H2O adj(ng/l)*10K):	785	487

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT
TEST

Form 106-18000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/20/2016 16:36

Test	AC	Time
01 Room Air	0.000	16:37
02 *Subject Test	REF*	16:37
03 Room Air	0.000	16:38

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = REFUSAL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

REFUSAL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/20/2016 16:39

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/20/2016 16:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:46
02 Std. Sol.	0.000	16:47
03 Room Air	0.000	16:47
04 Std. Sol.	0.000	16:48
05 Room Air	0.000	16:48
06 Std. Sol.	0.000	16:49
07 Room Air	0.000	16:50

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ZERO AC
0.000AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/20/2016 16:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:52
02 Std. Sol.	0.020	16:53
03 Room Air	0.000	16:53
04 Std. Sol.	0.020	16:54
05 Room Air	0.000	16:55
06 Std. Sol.	0.020	16:55
07 Room Air	0.000	16:56

08 Sim Temp = 34.0°C

Simul Ser No = MP3066
Std Sol No = 15160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/20/2016 16:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:58
02 Std. Sol.	0.394	16:59
03 Room Air	0.000	17:00
04 Std. Sol.	0.394	17:00
05 Room Air	0.000	17:01
06 Std. Sol.	0.396	17:02
07 Room Air	0.000	17:02

08 Sim Temp = 34.0°C

Simul Ser No = MP3068

Std Sol No = 14102

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:30

Test	AC	Time
01 Room Air	0.000	09:31
02 *Subject Test	INT*	09:31
03 Room Air	0.000	09:32

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10AC+0.05ACE
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: INTERFERENT DETECT
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:33

Test	AC	Time
01 Room Air	0.000	09:34
02 *Subject Test	INT*	09:35
03 Room Air	0.000	09:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.05 ACETONE
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECT*
0.05% ACETONE


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:37

Test	AC	Time
01 Room Air	0.000	09:38
02 *Subject Test	INT*	09:38
03 Room Air	0.000	09:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 METHANOL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECT*
0.10% METHANOL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:40

Test	AC	Time
01 Room Air	0.000	09:41
02 *Subject Test	INT*	09:41
03 Room Air	0.000	09:42

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 IPA
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECT*
0.10% ISOPROPANOL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:44

Test	AC	Time
01 Room Air	AMB*	09:45
02 Room Air	0.000	09:46

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

AMBIENT FAIL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:50

Test	AC	Time
01 Room Air	0.000	09:51
02 Invalid Sample	X.XXX	09:51
03 Room Air	0.000	09:52

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INVALID SAMPLE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:53

Test	AC	Time
01 Room Air	0.000	09:54
02 *Subject Test	ABT*	09:55
03 Room Air	0.000	09:56

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE ABORTED


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 09:57

Test	AC	Time
01 Room Air	0.000	09:58
02 Subject Test 1	0.000	09:58
03 Room Air	0.000	09:59
04 Reported AC	0.000	09:58

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA TEST
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 10:00

Test	AC	Time
01 Diagnostic	OK	10:01
02 Room Air	0.000	10:01
03 Subject Test 1	0.143	10:03
04 Room Air	0.000	10:05
05 Std. Gas	0.000*	10:07
06 Room Air	0.000	10:07

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL OUT OF TOL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK OUT
OF TOLERANCE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 10:10

Test	AC	Time
01 Diagnostic	OK	10:11
02 Room Air	0.000	10:11
03 Subject Test 1	0.142	10:12
04 Room Air	0.000	10:15
05 Std. Gas	0.081	10:16
06 Room Air	0.000	10:17
07 *Subject Test	RFI*	10:18
08 Room Air	0.000	10:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 10:26

Test	AC	Time
01 Diagnostic	OK	10:27
02 Room Air	0.000	10:28
03 *Subject Test	0.000*	10:31
04 Room Air	0.000	10:36
05 Std. Gas	0.081	10:37
06 Room Air	0.000	10:39
07 *Subject Test	0.000*	10:42
08 Room Air	0.000	10:44


No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DEFICIENT SAMPLES

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 10:46

Test	AC	Time
01 Diagnostic	OK	10:47
02 Room Air	IPS*	10:47
03 Room Air	0.000	10:48

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = IMPROPER SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

*IMPROPER
SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 12:14

Test	AC	Time
01 Diagnostic	OK	12:14
02 Room Air	0.000	12:15
03 Subject Test 1	0.138*	12:15
04 Room Air	0.000	12:18
05 Std. Gas	0.080	12:19
06 Room Air	0.000	12:20
07 Subject Test 2	0.192*	12:21
08 Room Air	0.000	12:22

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *DIFFERENCE
TOO GREAT*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 12:40


Test	AC	Time
01 Diagnostic	OK	12:41
02 Room Air	0.000	12:41
03 Subject Test 1	0.289	12:42
04 Room Air	0.000	12:45
05 Std. Gas	0.079	12:46
06 Room Air	0.000	12:48
07 Subject Test 2	0.294	12:48
08 Room Air	0.000	12:49
09 Reported AC	0.289	12:42

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS TEST
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

CMS TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = TOXL 8164.13.00 06/09
01/21/2016 12:57

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Gas	0.081	12:58
03 Room Air	0.000	12:58
04 Std. Gas	0.081	12:59
05 Room Air	0.000	12:59
06 Std. Gas	0.081	13:00
07 Room Air	0.000	13:00
08 Std. Gas	0.081	13:01
09 Room Air	0.000	13:01
10 Std. Gas	0.082	13:01
11 Room Air	0.000	13:02
12 Std. Gas	0.081	13:02
13 Room Air	0.000	13:03
14 Std. Gas	0.082	13:03
15 Room Air	0.000	13:04
16 Std. Gas	0.081	13:04
17 Room Air	0.000	13:05
18 Std. Gas	0.081	13:05
19 Room Air	0.000	13:06
20 Std. Gas	0.082	13:06
21 Room Air	0.000	13:07
22 Std. Gas	0.081	13:07
23 Room Air	0.000	13:07
24 Std. Gas	0.082	13:08
25 Room Air	0.000	13:08
26 Std. Gas	0.082	13:09
27 Room Air	0.000	13:09
28 Std. Gas	0.081	13:10
29 Room Air	0.000	13:10
30 Std. Gas	0.082	13:11
31 Room Air	0.000	13:11
32 Std. Gas	0.082	13:11
33 Room Air	0.000	13:12
34 Std. Gas	0.081	13:12
35 Room Air	0.000	13:13
36 Std. Gas	0.081	13:13
37 Room Air	0.000	13:14
38 Std. Gas	0.082	13:14
39 Room Air	0.000	13:15
40 Std. Gas	0.082	13:15
41 Room Air	0.000	13:16
42 Cal Check Stats:		
43 Average	0.0815	
44 Std Dev	0.0005	
45 Rel Std Dev	0.63%	

STABILITY
TEST
Charles J. Edwards

Operator Signature
CHARLES EDER

Remarks:

LOT: 29614080A2
EXP. DATE: 12/5/2016
CYL: 006

Form 106-I8000

STABILITY TEST

Charles Eder