

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006511

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5114 Lot No. 201505B Rep AC 0.030 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP306B Lot No. 14102 Rep AC 0.400 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS 4 Solution 0.10AC+0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A3 Cyl No. 028 Exp. Date 6.5.2016

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.080</u>	Test 5 <u>0.082</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 815 psi Reg ~800 psi

K. Remarks/Maintenance record: _____

4.1.2016

Date

Charles E. Edle

Field Inspector's Signature

01 Apr 2016

Date

Deb Shanaver

Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 12:53

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A3
Standard Cyl #? 28
Standard Expiration? 06/05/2016
Oper No? 666666

Flow Cal. Date: 08/21/2015
Slope 688
Intercept -714727

IR Calibration Date: 08/21/2015

	3um	9um
0th Coef(*100):	-19744	-15946
1st Coef(*100):	272125	142950
2nd Coef(*100):	2975	1383
H2O adj(mg/l*10k):	725	394

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 12:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:58
02 Std. Sol.	0.027	12:59
03 Room Air	0.000	13:00
04 Std. Sol.	0.028	13:01
05 Room Air	0.000	13:01
06 Std. Sol.	0.029	13:02
07 Room Air	0.000	13:03

08 Sim Temp = 34.0°C

Simul Ser No = DR5114
Std Sol No = 201505B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.030 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 13:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:05
02 Std. Sol.	0.406	13:06
03 Room Air	0.000	13:06
04 Std. Sol.	0.411	13:07
05 Room Air	0.000	13:08
06 Std. Sol.	0.410	13:09
07 Room Air	0.000	13:10

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
 0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 13:13

Test	AC	Time
01 Room Air	0.000	13:14
02 *Subject Test	INT*	13:15
03 Room Air	0.000	13:16

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 13:18

Test	AC	Time
01 Diagnostic	OK	13:19
02 Room Air	0.000	13:19
03 *Subject Test	RFI*	13:19
04 Room Air	0.000	13:20

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 28
Expiration Date = 06/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 13:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:22
02 Std. Gas	0.081	13:22
03 Room Air	0.000	13:23
04 Std. Gas	0.080	13:23
05 Room Air	0.000	13:24
06 Std. Gas	0.080	13:24
07 Room Air	0.000	13:25

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 13:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:26
02 Std. Gas	0.081	13:26
03 Room Air	0.000	13:27
04 Std. Gas	0.082	13:27
05 Room Air	0.000	13:28
06 Std. Gas	0.081	13:28
07 Room Air	0.000	13:28

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006511
Location = TOXL 8164.13.00 06/09
04/01/2016 13:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.082	13:30
03 Room Air	0.000	13:30
04 Std. Gas	0.081	13:31
05 Room Air	0.000	13:31
06 Std. Gas	0.081	13:31
07 Room Air	0.000	13:32

Lot No = 13314080A3
Cyl No = 28
Exp Date = 06/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC