

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006513

Location: TOXL

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5114 Lot No. 201505B Rep AC 0.030 ✓
- F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC+0.05%ACE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 13314080A3 Cyl No. 028 Exp. Date 6/5/2016

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 851 psi Reg 2850 psi

K. Remarks/Maintenance record: \_\_\_\_\_

4.1.2016  
Date

[Signature]  
Field Inspector's Signature

01 Apr 2016  
Date

[Signature]  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006513  
Location = TOXL 8164.13.00 06/09  
04/01/2016 10:59

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstu vwxyz1234567890- =|  
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &\* () \_ +?

abcdefghijklmnopqrstu vwxyz1234567890- =|  
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^ &\* () \_ +?

Current Instrument Setup


Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 13314080A3  
Standard Cyl #? 28  
Standard Expiration? 06/05/2016  
Oper No? 666666

Flow Cal. Date: 08/20/2015  
Slope 680  
Intercept -696037

IR Calibration Date: 08/21/2015

	3um	9um
0th Coef(*100):	-14436	-13610
1st Coef(*100):	265490	136147
2nd Coef(*100):	4422	1710
H2O adj (mg/l*10k):	672	448

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

**PRINT TEST**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Sol.	0.028	11:02
03 Room Air	0.000	11:03
04 Std. Sol.	0.029	11:03
05 Room Air	0.000	11:04
06 Std. Sol.	0.029	11:05
07 Room Air	0.000	11:05

08 Sim Temp = 34.0°C

Simul Ser No = DR5114  
Std Sol No = 201505B  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.030 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Sol.	0.404	11:08
03 Room Air	0.000	11:08
04 Std. Sol.	0.406	11:09
05 Room Air	0.000	11:10
06 Std. Sol.	0.406	11:10
07 Room Air	0.000	11:11

08 Sim Temp = 34.0°C

Simul Ser No = MP3068  
Std Sol No = 14102  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      *HIGH AC*  
                    *0.400 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:13

Test	AC	Time
01 Room Air	0.000	11:14
02 *Subject Test	INT*	11:14
03 Room Air	0.000	11:15

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 13314080A3  
Cyl No = 28  
Expiration Date = 06/05/2016  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT CHECK*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:16

Test	AC	Time
01 Diagnostic	OK	11:17
02 Room Air	0.000	11:18
03 *Subject Test	RFI*	11:18
04 Room Air	0.000	11:19

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 13314080A3  
Cyl No = 28  
Expiration Date = 06/05/2016  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Gas	0.080	11:21
03 Room Air	0.000	11:21
04 Std. Gas	0.080	11:21
05 Room Air	0.000	11:22
06 Std. Gas	0.081	11:22
07 Room Air	0.000	11:23

Lot No = 13314080A3  
Cyl No = 28  
Exp Date = 06/05/2016  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:24
02 Std. Gas	0.081	11:24
03 Room Air	0.000	11:25
04 Std. Gas	0.081	11:25
05 Room Air	0.000	11:25
06 Std. Gas	0.081	11:26
07 Room Air	0.000	11:26

Lot No = 13314080A3  
Cyl No = 28  
Exp Date = 06/05/2016  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006513  
Location = TOXL      8164.13.00 06/09  
04/01/2016      11:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:27
02 Std. Gas	0.081	11:28
03 Room Air	0.000	11:28
04 Std. Gas	0.081	11:28
05 Room Air	0.000	11:29
06 Std. Gas	0.081	11:29
07 Room Air	0.000	11:30

Lot No = 13314080A3  
Cyl No = 28  
Exp Date = 06/05/2016  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080Ac