



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (03-2014)

Serial Number 80-006510	Instrument Location HETT
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

### Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): NA
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>J Collins</i>	Date <i>05/13/2016</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>20 May 16</i>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Charles E. Edm* Certified Date: *18 Oct. 16*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.13.00 06/09  
05/13/2016      10:07

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      29614080A2  
Standard Cyl #?      34  
Standard Expiration?      12/05/2016  
Oper No?      020533

Flow Cal. Date:      08/24/2015  
Slope      624  
Intercept      -539297

IR Calibration Date:      08/24/2015

	3um	9um
0th Coef(*100):	-23613	-13697
1st Coef(*100):	251959	131614
2nd Coef(*100):	4659	1853
H2O adj(mg/l*10k):	819	493

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
TRAVIS COLLINS

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.13.00 06/09  
05/13/2016      10:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:09
02 Std. Gas	0.079	10:09
03 Room Air	0.000	10:09
04 Std. Gas	0.080	10:10
05 Room Air	0.000	10:10
06 Std. Gas	0.080	10:11
07 Room Air	0.000	10:11

Lot No = 29614080A2  
Cyl No = 34  
Exp Date = 12/05/2016  
County = 01      Oper No. = 020533



Operator Signature  
TRAVIS COLLINS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006510  
Location = HETT      8164.13.00 06/09  
05/13/2016      10:14

Test	AC	Time
01 Room Air	0.000	10:17
02 *Subject Test	RFI*	10:17
03 Room Air	0.000	10:17

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, JOE T  
Sub DOB = 10/10/2014  
Sub Sex = Male      Weight = 188  
Test = MIC      Cit = 012345  
Dr. Lic. = ND/012345  
Lot No = 29614080A2  
Cyl No = 34  
Expiration Date = 12/05/2016  
County = 01      Oper No. = 020533



Operator Signature  
TRAVIS COLLINS

Remarks: