



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number <i>80-006508</i>	Instrument Location <i>Richland County</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>5-17-16 ✓</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>20 May 16</i>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *[Signature]* Certified Date: *10 Nov. 16*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006508
Location = WAHP 8164.13.00 06/09
05/17/2016 16:04

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNopqrstUVWXYz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMNopqrstUVWXYz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01316080A2
Standard Cyl #? 34
Standard Expiration? 02/05/2018
Oper No? 020401

Flow Cal. Date: 08/20/2015
Slope 644
Intercept -568181

IR Calibration Date: 08/20/2015
 3um 9um

0th Coef(*100): -22365 -18627
1st Coef(*100): 256969 136095
2nd Coef(*100): 3168 1444
H2O adj(mg/l*10k): 693 453

***** Printer Test End *****



Operator Signature
MATTHEW ANDERSON

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006508
Location = WAHP 8164.13.00 06/09
05/17/2016 16:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:06
02 Std. Gas	0.078	16:06
03 Room Air	0.000	16:07
04 Std. Gas	0.079	16:07
05 Room Air	0.000	16:08
06 Std. Gas	0.079	16:08
07 Room Air	0.000	16:09

Lot No = 01316080A2

Cyl No = 34

Exp Date = 02/05/2018

County = 39

Oper No. = 020401



Operator Signature
MATTHEW ANDERSON

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006508
Location = WAHP 8164.13.00 06/09
05/17/2016 16:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	16:10
02 Room Air	0.000	16:11

*Invalid Test
Inhibited - RFI

Lot No = 01316080A2
Cyl No = 34
Exp Date = 02/05/2018
County = 39 Oper No. = 020401



Operator Signature
MATTHEW ANDERSON

Remarks:

Form 106-I8000