

**INTOXILYZER® 8000 INITIAL INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-006669

**A. GENERAL Setup and Checks:**

1.  Power On
  - 5 min. countdown during initial warm-up
  - 1 min. countdown to begin test after standby mode
2.  Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
  - Level 1 (Operator)    Level 2 (Field Inspector)    Level 3 (Lab Personnel)
4.  Printer Setup (Level 3,S,P)
5.  Forms Load (CMI, Inc., CoBrA Software)

**B. LEVEL 1 Setup and Checks:**

1.  Internal Printer Test and Paper Feed (Level 1,P). Attach record.
2.  External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
  - Operator ID    ND License    R key: Refusal. Attach record.

**C. LEVEL 2 Setup and Checks:**

1.  Continuous Air Blank (Level 2,A)
2.  Diagnostic (Level 2,D). Attach record.
3.  Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
  - Inst Ser#: 80-006669   DSP Ver: 132 1362.14
  - Core Ver: 8164.14   CPLD Ver: 7

**D. LEVEL 3 Setup and Checks:**

1. DVM Monitor (Level 3,D,D): 3: 12944 s: 1 C: 47.00  
9: 13147 s: 1 B: 45. 0  
(RFI indicator with keyed radio): 1
2. Function (F) Keys: :  F5 – Tone                       F9 – Fan  
 F6 – Green LED                       F10 – Pump  
 F7 – Red LED                           F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 962 CF: 1.05
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 196 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0  
B: 45.1 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 161 psi P (gauge): 175 psi
7.  General Setup (Level 3,S,G)
8.  Configure Start Test (Level 3,S,T)
9.  Network (Level 3,S,N)

**E. TESTS (attach all test records):**

**Configure wet bath simulator standards before running tests (Level 1,S):**

1. Zero AC (Level 1,C):  
Sim ser #: MP3064 Lot #: MILLI Q H<sub>2</sub>O Exp date: NA  
Known value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):  
Sim ser #: MP3003 Lot #: 2016 0314 Exp date: 3.16.2018  
Known value: 0.010 AC  
Results: 0.011 AC 0.011 AC 0.011 AC
3. High AC (Level 1,C):  
Sim ser #: MP3063 Lot #: 16102 Exp date: 3.22.18  
Known value: 0.400 AC  
Results: 0.394 AC 0.396 AC 0.396 AC

**Configure dry gas standard before running tests (Level 1,S):**

4. Ethanol + 0.05% Acetone (Level 1,B):  
Sim ser #: DR5135 Lot #: ICS4 Exp date: 9/9/2017  
Known value: 0.10 AC + 0.05% ACETONE  
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):  
Sim ser #: DR5143 Lot #: ACE2 Exp date: 9/29/17  
Known value: 0.05% ACETONE  
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):  
Sim ser #: DR3377 Lot #: MET2 Exp date: 9/29/17  
Known value: 0.10% METHANOL  
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):  
Sim ser #: DR3453 Lot #: IPA2 Exp date: 9/29/17  
Known value: 0.10% ISOPROPANOL  
 Display reads "Interferent Detect"

- 8.  Ambient Fail (Level 1,B): Display and test record read correct.
- 9.  Invalid Sample (Level 1,B): Display and test record read correct.
- 10.  Sequence Aborted (Level 1,B): Display and test record read correct.
- 11.  ABA (Level 1,B): Test record reads correct.
- 12.  Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13.  RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14.  Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15.  Deficient Samples (CMS mode): Display and test record read correct.
- 16.  Improper Sample (CMS mode): Display and test record read correct.
- 17.  Difference Too Great (CMS mode): Display and test record read correct.
- 18.  CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0813 Std Dev: 0.0008 Rel Std Dev: 0.97%

Maintenance/Repairs/Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Instrument is acceptable to be used in the field.

Charles E. Ehr  
Field Inspector's Signature

1/10/2017  
Date

Debs Shanaver  
Reviewed by

13 Jan '17  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006669  
Location = TOXL 8164.14.00 09/16  
01/06/2017 16:29

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMNPO

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNPOQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNPOQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNPOQRSTUWXYZ123456


42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNPOQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup  
Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 19615080A5  
Standard Cyl #? 26  
Standard Expiration? 09/05/2017  
Oper No? 666666

Flow Cal. Date: 08/15/2016  
Slope 678  
Intercept -615536

IR Calibration Date: 08/15/2016  
3um 9um  
-----  
0th Coef(\*100): -19192 -25323  
1st Coef(\*100): 259438 128900  
2nd Coef(\*100): 2747 1289  
H2O adj(mg/l\*10K): 613 563

\*\*\*\*\* Test End \*\*\*\*\*

  
Operator Signature  
CHARLES EDER

Remarks:  
*Internal Print*

Form 106-18000

*TEST*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                 8164.14.00 09/16  
01/06/2017                        16:30

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:                Enabled  
Start Test Sequence:            DABACABA  
Display Prelim Rslt?            Yes  
Display Third Digit?            Yes  
Inhib Printer(Y/N)?            No  
Display Volume?                No  
Disable On Memfull?            No  
# of Print Copies?              1  
Select Std (D/W/I)?            Dry  
Standard Value?                0.080  
Standard Lot #?                 80  
Standard Cyl #?                765432  
Standard Expiration?           11/11/2016  
Oper No?                        666666

Flow Cal. Date:                 08/15/2016  
Slope                            678  
Intercept                       -615536

IR Calibration Date:            08/15/2016  
                                  3um                            9um

-----  
0th Coef(\*100):                -19192                        -25323  
1st Coef(\*100):                259438                        128900  
2nd Coef(\*100):                2747                            1289  
H2O adj(mg/l\*10k):            613                            563

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

EXTERNAL PRINT  
TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006669  
Location = TOXL      8164.14.00 09/16  
01/06/2017      16:32

Test	AC	Time
01 Room Air	0.000	16:33
02 *Subject Test	REF*	16:33
03 Room Air	0.000	16:34

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = R KEY REFUSAL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: "R" KEY  
REFUSAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                 8164.14.00 09/16  
01/06/2017                        16:35

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

DIAGNOSTIC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/06/2017                            16:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:43
02 Std. Sol.	0.000	16:44
03 Room Air	0.000	16:45
04 Std. Sol.	0.000	16:45
05 Room Air	0.000	16:46
06 Std. Sol.	0.000	16:46
07 Room Air	0.000	16:47

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = MILLIQ H2O  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

ZERO AC  
0.000 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                8164.14.00 09/16  
01/09/2017                        12:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Sol.	0.011	12:57
03 Room Air	0.000	12:58
04 Std. Sol.	0.011	12:59
05 Room Air	0.000	12:59
06 Std. Sol.	0.011	13:00
07 Room Air	0.000	13:00

08 Sim Temp = 34.0°C

Simul Ser No = MP3003

Std Sol No = 201603H

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.010 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006669  
Location = TOXL      8164.14.00 09/16  
01/09/2017      13:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:06
02 Std. Sol.	0.394	13:07
03 Room Air	0.000	13:07
04 Std. Sol.	0.396	13:08
05 Room Air	0.000	13:09
06 Std. Sol.	0.396	13:09
07 Room Air	0.000	13:10

08 Sim Temp = 34.0°C

Simul Ser No = MP3063

Std Sol No = 16102

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/09/2017                            13:12

Test	AC	Time
01 Room Air	0.000	13:13
02 *Subject Test	INT*	13:13
03 Room Air	0.000	13:14

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10AC+0.05ACE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.10 AC + 0.05% ACETONE*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/09/2017                            13:15

Test	AC	Time
01 Room Air	0.000	13:16
02 *Subject Test	INT*	13:17
03 Room Air	0.000	13:17

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.05 ACETONE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.05% ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/09/2017                            13:18

Test	AC	Time
01 Room Air	0.000	13:19
02 *Subject Test	INT*	13:19
03 Room Air	0.000	13:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10 METHANOL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.10% METANOL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006669  
Location = TOXL      8164.14.00 09/16  
01/09/2017      13:21

Test	AC	Time
01 Room Air	0.000	13:21
02 *Subject Test	INT*	13:22
03 Room Air	0.000	13:22

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10 IPA  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.10% ISOPROPNOL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/09/2017                            13:34

Test	AC	Time
01 Room Air	AMB*	13:34
02 Room Air	0.000	13:35

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*AMBIENT FAIL*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                8164.14.00 09/16  
01/09/2017                        13:36

Test	AC	Time
01 Room Air	0.000	13:37
02 Invalid Sample	X.XXX	13:37
03 Room Air	0.000	13:38

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = INVALID SAMP  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*INVALID SAMPLE*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                 8164.14.00 09/16  
01/09/2017                        13:38

Test	AC	Time
01 Room Air	0.000	13:39
02 *Subject Test	ABT*	13:39
03 Room Air	0.000	13:40

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                 Weight = 150  
Test = DUI                         Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                         Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: SEQUENCE ABORTED

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006669  
Location = TOXL                8164.14.00 09/16  
01/09/2017                      13:41

Test	AC	Time
01 Room Air	0.000	13:41
02 Subject Test 1	0.276	13:42
03 Room Air	0.000	13:43
04 Reported AC	0.276	13:42

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = ABA MODE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *SINGLE BREATH TEST*  
*ABA MODE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/09/2017                            13:43

Test	AC	Time
01 Diagnostic	OK	13:44
02 Room Air	0.000	13:44
03 Subject Test 1	0.275	13:45
04 Room Air	0.000	13:47
05 Std. Gas	0.091*	13:48
06 Room Air	0.000	13:49

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = CAL OUT TOL  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK OUT  
OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/10/2017                            10:38

Test	AC	Time
01 Diagnostic	OK	10:39
02 Room Air	0.000	10:40
03 Subject Test 1	0.275	10:40
04 Room Air	0.000	10:43
05 Std. Gas	0.082	10:44
06 Room Air	0.000	10:45
07 *Subject Test	RFI*	10:45
08 Room Air	0.000	10:46

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI DETECTED  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:    RFI DETECTED

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/10/2017                            10:47

Test	AC	Time
01 Diagnostic	OK	10:48
02 Room Air	0.000	10:48
03 *Subject Test	0.000*	10:51
04 Room Air	0.000	10:54
05 Std. Gas	0.080	10:55
06 Room Air	0.000	10:56
07 *Subject Test	0.000*	10:59
08 Room Air	0.000	11:00

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DEFICIENT SAMP  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: DEFICIENT SAMPLES

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/10/2017                            11:06

Test	AC	Time
01 Diagnostic	OK	11:07
02 Room Air	0.000	11:07
03 Subject Test 1	0.274	11:08
04 Room Air	0.000	11:10
05 Std. Gas	0.081	11:11
06 Room Air	IPS*	11:12
07 Room Air	0.000	11:12

\*Improper Sample

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = IMPROPER SAMP  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*Improper Sample*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/10/2017                            11:13

Test	AC	Time
01 Diagnostic	OK	11:14
02 Room Air	0.000	11:15
03 Subject Test 1	0.274*	11:15
04 Room Air	0.000	11:17
05 Std. Gas	0.081	11:19
06 Room Air	0.000	11:20
07 Subject Test 2	0.210*	11:20
08 Room Air	0.000	11:21

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

DIFFERENCE  
TOO GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/10/2017                            11:22

Test	AC	Time
01 Diagnostic	OK	11:23
02 Room Air	0.000	11:23
03 Subject Test 1	0.271	11:23
04 Room Air	0.000	11:26
05 Std. Gas	0.080	11:27
06 Room Air	0.000	11:28
07 Subject Test 2	0.269	11:29
08 Room Air	0.000	11:30
09 Reported AC	0.269	11:29

Difference OK  
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = CMS MODE  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 26  
Expiration Date = 09/05/2017  
County = 08                            Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: CMS MODE

EVIDENTIAL TEST

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006669  
Location = TOXL                    8164.14.00 09/16  
01/10/2017                            11:31

Test	AC	Time
01 Room Air	0.000	11:32
02 Std. Gas	0.079	11:32
03 Room Air	0.000	11:32
04 Std. Gas	0.080	11:33
05 Room Air	0.000	11:33
06 Std. Gas	0.081	11:34
07 Room Air	0.000	11:34
08 Std. Gas	0.081	11:34
09 Room Air	0.000	11:35
10 Std. Gas	0.082	11:35
11 Room Air	0.000	11:36
12 Std. Gas	0.081	11:36
13 Room Air	0.000	11:37
14 Std. Gas	0.081	11:37
15 Room Air	0.000	11:38
16 Std. Gas	0.081	11:38
17 Room Air	0.000	11:39
18 Std. Gas	0.081	11:39
19 Room Air	0.000	11:39
20 Std. Gas	0.082	11:40
21 Room Air	0.000	11:40
22 Std. Gas	0.082	11:41
23 Room Air	0.000	11:41
24 Std. Gas	0.082	11:41
25 Room Air	0.000	11:42
26 Std. Gas	0.082	11:42
27 Room Air	0.000	11:43
28 Std. Gas	0.081	11:43
29 Room Air	0.000	11:44
30 Std. Gas	0.082	11:44
31 Room Air	0.000	11:45
32 Std. Gas	0.081	11:45
33 Room Air	0.000	11:45
34 Std. Gas	0.081	11:46
35 Room Air	0.000	11:46
36 Std. Gas	0.082	11:47
37 Room Air	0.000	11:47
38 Std. Gas	0.082	11:48
39 Room Air	0.000	11:48
40 Std. Gas	0.081	11:48
41 Room Air	0.000	11:49
42 Cal Check Stats:		
43 Average	0.0813	
44 Std Dev	0.0008	
45 Rel Std Dev	0.97%	



Operator Signature  
CHARLES EDER

Remarks: STABILITY TEST

Form 106-I8000

Lot: 19615080A5

C4L: 026

Exp: 9/5/2017