

**INTOXILYZER® 8000 INITIAL INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-006676

**A. GENERAL Setup and Checks:**

1.  Power On
  - 5 min. countdown during initial warm-up
  - 1 min. countdown to begin test after standby mode
2.  Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
  - Level 1 (Operator)     Level 2 (Field Inspector)     Level 3 (Lab Personnel)
4.  Printer Setup (Level 3,S,P)
5.  Forms Load (CMI, Inc., CoBrA Software)

**B. LEVEL 1 Setup and Checks:**

1.  Internal Printer Test and Paper Feed (Level 1,P). Attach record.
2.  External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
  - Operator ID     ND License     R key: Refusal. Attach record.

**C. LEVEL 2 Setup and Checks:**

1.  Continuous Air Blank (Level 2,A)
2.  Diagnostic (Level 2,D). Attach record.
3.  Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
  - Inst Ser#: 80-006676 DSP Ver: 1362.14
  - Core Ver: 8164.14 CPLD Ver: 7

**D. LEVEL 3 Setup and Checks:**

1. DVM Monitor (Level 3,D,D): 3: 12763 s: 0 C: 47.00  
9: 13292 s: 1 B: 45 0  
(RFI indicator with keyed radio): L
2. Function (F) Keys: :  F5 – Tone                             F9 - Fan  
 F6 – Green LED                         F10 – Pump  
 F7 – Red LED                             F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 943 CF: 1.07
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 222 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.5 I: 0.0  
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 882 psi P (gauge): ~875 psi
7.  General Setup (Level 3,S,G)
8.  Configure Start Test (Level 3,S,T)
9.  Network (Level 3,S,N)

**E. TESTS (attach all test records):**

**Configure wet bath simulator standards before running tests (Level 1,S):**

1. Zero AC (Level 1,C):  
Sim ser #: MP3064 Lot #: Milli H<sub>2</sub>O Exp date: N/A  
Known value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):  
Sim ser #: MP3003 Lot #: 20160314 Exp date: 3.16.2018  
Known value: 0.010 AC  
Results: 0.010 AC 0.009 AC 0.009 AC
3. High AC (Level 1,C):  
Sim ser #: MP3063 Lot #: 16102 Exp date: 3.22.18  
Known value: 0.400 AC  
Results: 0.393 AC 0.395 AC 0.395 AC

**Configure dry gas standard before running tests (Level 1,S):**

4. Ethanol + 0.05% Acetone (Level 1,B):  
Sim ser #: DR5135 Lot #: ICS4 Exp date: 9.9.2017  
Known value: 0.10 AC + 0.05% ACETONE  
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):  
Sim ser #: DR5143 Lot #: ACE2 Exp date: 9.29.2017  
Known value: 0.05% ACETONE  
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):  
Sim ser #: DR3377 Lot #: MET2 Exp date: 9.29.2017  
Known value: 0.10% METHANOL  
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):  
Sim ser #: DR3453 Lot #: IPA2 Exp date: 9.29.2017  
Known value: 0.10% IPA (ISOPROPANOL)  
 Display reads "Interferent Detect"

- 8.  Ambient Fail (Level 1,B): Display and test record read correct.
- 9.  Invalid Sample (Level 1,B): Display and test record read correct.
- 10.  Sequence Aborted (Level 1,B): Display and test record read correct.
- 11.  ABA (Level 1,B): Test record reads correct.
- 12.  Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13.  RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14.  Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15.  Deficient Samples (CMS mode): Display and test record read correct.
- 16.  Improper Sample (CMS mode): Display and test record read correct.
- 17.  Difference Too Great (CMS mode): Display and test record read correct.
- 18.  CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0825 Std Dev: 0.0008 Rel Std Dev: 0.93%

Maintenance/Repairs/Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Instrument is acceptable to be used in the field.

  
\_\_\_\_\_  
Field Inspector's Signature

<sup>19</sup>  
~~1/18/17~~  
1/18/17  
\_\_\_\_\_  
Date

  
\_\_\_\_\_  
Reviewed by

19 Jan 17  
\_\_\_\_\_  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/18/2017      16:25

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

32 characters per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

42 characters per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

Screen Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Menfull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 80  
Standard Cyl #? 1234567  
Standard Expiration? 11/11/2017  
Oper No? 666666

Flow Cal. Date: 08/22/2016  
Slope 710  
Intercept -713681

IR Calibration Date: 08/22/2016  
3um 9um

0th Coef(*100):	-16667	-22327
1st Coef(*100):	259190	132291
2nd Coef(*100):	3393	1374
H2O adj(mg/l*10K):	611	546

\*\*\*\*\* Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: *Internal  
PRINT  
TEST*  
Form 106-18000





Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/18/2017      16:42

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

DIAGNOSTIC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/18/2017      16:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:48
02 Std. Sol.	0.000	16:49
03 Room Air	0.000	16:50
04 Std. Sol.	0.000	16:50
05 Room Air	0.000	16:51
06 Std. Sol.	0.000	16:52
07 Room Air	0.000	16:52

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = MILLIQ H2O  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

ZERO AC  
0.000 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      12:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:35
02 Std. Sol.	0.010	12:36
03 Room Air	0.000	12:37
04 Std. Sol.	0.009	12:37
05 Room Air	0.000	12:38
06 Std. Sol.	0.009	12:39
07 Room Air	0.000	12:39

08 Sim Temp = 34.0°C

Simul Ser No = MP3003  
Std Sol No = 201603H  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.010 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      12:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:41
02 Std. Sol.	0.393	12:42
03 Room Air	0.000	12:42
04 Std. Sol.	0.395	12:43
05 Room Air	0.000	12:44
06 Std. Sol.	0.395	12:44
07 Room Air	0.000	12:45


08 Sim Temp = 34.0°C

Simul Ser No = MP3063

Std Sol No = 16102

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      12:47

Test	AC	Time
01 Room Air	0.000	12:48
02 *Subject Test	INT*	12:48
03 Room Air	0.000	12:49

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10AC+0.05ACE  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*

*0.10 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      12:49

Test	AC	Time
01 Room Air	0.000	12:50
02 *Subject Test	INT*	12:50
03 Room Air	0.000	12:51

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.05 ACETONE  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.05% ACETONE*  
Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
01/19/2017                            12:52

Test	AC	Time
01 Room Air	0.000	12:53
02 *Subject Test	INT*	12:53
03 Room Air	0.000	12:54

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10 METHANOL  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*

*0.10% METHANOL*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      12:54

Test	AC	Time
01 Room Air	0.000	12:55
02 *Subject Test	INT*	12:56
03 Room Air	0.000	12:57

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10 IPA  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*

*0.10% ISOPROPA<sub>2</sub>NOL*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
01/19/2017                            12:57

Test	AC	Time
01 Room Air	AMB*	12:58
02 Room Air	0.000	12:59

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

AMBIENT FAIL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006676  
Location = TOXL                8164.14.00 09/16  
01/19/2017                        12:59

Test	AC	Time
01 Room Air	0.000	13:00
02 Invalid Sample	X.XXX	13:00
03 Room Air	0.000	13:01

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                Weight = 150  
Test = DUI                      Cit = INVALID SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08                      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INVALID SAMPLE*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      13:01

Test	AC	Time
01 Room Air	0.000	13:02
02 *Subject Test	ABT*	13:02
03 Room Air	0.000	13:03

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: SEQUENCE ABORTED

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      13:03

Test	AC	Time
01 Room Air	0.000	13:04
02 Subject Test 1	0.035	13:04
03 Room Air	0.000	13:05
04 Reported AC	0.035	13:04

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = ABA MODE  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *ABA MODE*

*SINGLE BREATH TEST*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      13:05

Test	AC	Time
01 Diagnostic	OK	13:06
02 Room Air	0.000	13:07
03 Subject Test 1	0.265	13:07
04 Room Air	0.000	13:09
05 Std. Gas	0.091*	13:10
06 Room Air	0.000	13:11

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = CAL OUT OF TOL  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
OUT OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      13:29

Test	AC	Time
01 Diagnostic	OK	13:30
02 Room Air	0.000	13:30
03 Subject Test 1	0.265	13:31
04 Room Air	0.000	13:33
05 Std. Gas	0.082	13:34
06 Room Air	0.000	13:36
07 Reference	RFI*	13:36
08 Room Air	0.000	13:36

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI DETECTED  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI DETECTED*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006676  
Location = TOXL                    8164.14.00 09/16  
01/19/2017                            13:39

Test	AC	Time
01 Diagnostic	OK	13:40
02 Room Air	0.000	13:40
03 *Subject Test	0.000*	13:43
04 Room Air	0.000	13:46
05 Std. Gas	0.082	13:47
06 Room Air	0.000	13:48
07 *Subject Test	0.000*	13:51
08 Room Air	0.000	13:52

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DEFICIENT SAMP  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08                            Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: DEFICIENT SAMPLES

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      13:53

Test	AC	Time
01 Diagnostic	OK	13:53
02 Room Air	IPS*	13:54
03 Room Air	0.000	13:54

\*Improper Sample

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = IMPROPER SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*IMPROPER  
SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      13:54

Test	AC	Time
01 Diagnostic	OK	13:56
02 Room Air	0.000	13:57
03 Subject Test 1	0.264*	13:57
04 Room Air	0.000	13:59
05 Std. Gas	0.083	14:01
06 Room Air	0.000	14:02
07 Subject Test 2	0.207*	14:02
08 Room Air	0.000	14:03

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: DIFFERENCE  
Too GREAT

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006676  
Location = TOXL      8164.14.00 09/16  
01/19/2017      14:06


Test	AC	Time
01 Diagnostic	OK	14:07
02 Room Air	0.000	14:07
03 Subject Test 1	0.261	14:08
04 Room Air	0.000	14:10
05 Std. Gas	0.082	14:11
06 Room Air	0.000	14:13
07 Subject Test 2	0.260	14:13
08 Room Air	0.000	14:14
09 Reported AC	0.260	14:13

Difference OK  
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = CMS MODE  
Dr. Lic. = ND/DIS821456  
Lot No = 09216080A3  
Cyl No = 38  
Expiration Date = 05/05/2018  
County = 08      Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: CMS MODE

EVIDENTIAL BREATH TEST

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006676  
Location = TOXL                 8164.14.00 09/16  
01/19/2017                        14:16

Test	AC	Time
01 Room Air	0.000	14:17
02 Std. Gas	0.081	14:17
03 Room Air	0.000	14:18
04 Std. Gas	0.082	14:18
05 Room Air	0.000	14:18
06 Std. Gas	0.081	14:19
07 Room Air	0.000	14:19
08 Std. Gas	0.082	14:20
09 Room Air	0.000	14:20
10 Std. Gas	0.082	14:20
11 Room Air	0.000	14:21
12 Std. Gas	0.083	14:21
13 Room Air	0.000	14:22
14 Std. Gas	0.083	14:22
15 Room Air	0.000	14:23
16 Std. Gas	0.082	14:23
17 Room Air	0.000	14:24
18 Std. Gas	0.083	14:24
19 Room Air	0.000	14:24
20 Std. Gas	0.083	14:25
21 Room Air	0.000	14:25
22 Std. Gas	0.083	14:26
23 Room Air	0.000	14:26
24 Std. Gas	0.083	14:26
25 Room Air	0.000	14:27
26 Std. Gas	0.083	14:27
27 Room Air	0.000	14:28
28 Std. Gas	0.082	14:28
29 Room Air	0.000	14:29
30 Std. Gas	0.082	14:29
31 Room Air	0.000	14:30
32 Std. Gas	0.083	14:30
33 Room Air	0.000	14:30
34 Std. Gas	0.082	14:31
35 Room Air	0.000	14:31
36 Std. Gas	0.084	14:32
37 Room Air	0.000	14:32
38 Std. Gas	0.082	14:33
39 Room Air	0.000	14:33
40 Std. Gas	0.083	14:33
41 Room Air	0.000	14:34
42 Cal Check Stats:		
43 Average	0.0825	
44 Std Dev	0.0008	
45 Rel Std Dev	0.93%	



Operator Signature  
CHARLES EDER

Remarks: STABILITY  
TEST

Form 106-I8000

LOT: 0921608043

EXP: 5/5/2018

CYL: 038