

### INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006688

#### A. GENERAL Setup and Checks:

- Power On  
 5 min. countdown during initial warm-up  
 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
- Password Security:  
 Level 1 (Operator)    Level 2 (Field Inspector)    Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

#### B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
- Barcode Reader (Level 1,B):  
 Operator ID    ND License    R key: Refusal. Attach record.

#### C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
- Ser Number/Version Info (Level 2,V):  
Inst Ser#: 80-006688   DSP Ver: CEE  
Core Ver: 8164.14   CPLD Ver: 132 1362.14  
7

#### D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 13035 s: 0 c: 47.06  
9: 13338 s: 1 B: 45 0  
(RFI indicator with keyed radio): 1
- Function (F) Keys: :  F5 – Tone                       F9 - Fan  
 F6 – Green LED                       F10 – Pump  
 F7 – Red LED                          F11 – Solenoid
- Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 954 CF: 1.06
- Flow Monitor (Level 3,D,F): Flow Sensor, R: 239 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0  
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 711 psi P (gauge): ~725 psi
7.  General Setup (Level 3,S,G)
8.  Configure Start Test (Level 3,S,T)
9.  Network (Level 3,S,N)

**E. TESTS (attach all test records):**

**Configure wet bath simulator standards before running tests (Level 1,S):**

1. Zero AC (Level 1,C):  
Sim ser #: MP3064 Lot #: MILLI Q H<sub>2</sub>O Exp date: NA  
Known value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):  
Sim ser #: MP3060 Lot #: 201507F Exp date: 7.28.2017  
Known value: 0.020 AC  
Results: 0.019 AC 0.019 AC 0.019 AC
3. High AC (Level 1,C):  
Sim ser #: MP3063 Lot #: 16102 Exp date: 3/22/18  
Known value: 0.400 AC  
Results: 0.391 AC 0.392 AC 0.391 AC

**Configure dry gas standard before running tests (Level 1,S):**

4. Ethanol + 0.05% Acetone (Level 1,B):  
Sim ser #: DR5135 Lot #: ICS4 Exp date: 9/9/2017  
Known value: 0.10 AC + 0.05% ACETONE  
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):  
Sim ser #: DR5143 Lot #: ACE2 Exp date: 9/29/2017  
Known value: 0.05% ACETONE  
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):  
Sim ser #: DR3377 Lot #: MET2 Exp date: 9/29/2017  
Known value: 0.10% METHANOL  
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):  
Sim ser #: DR3453 Lot #: IPA2 Exp date: 9/29/2017  
Known value: 0.10% ISOPROPANOL  
 Display reads "Interferent Detect"

- 8.  Ambient Fail (Level 1,B): Display and test record read correct.
- 9.  Invalid Sample (Level 1,B): Display and test record read correct.
- 10.  Sequence Aborted (Level 1,B): Display and test record read correct.
- 11.  ABA (Level 1,B): Test record reads correct.
- 12.  Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13.  RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14.  Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15.  Deficient Samples (CMS mode): Display and test record read correct.
- 16.  Improper Sample (CMS mode): Display and test record read correct.
- 17.  Difference Too Great (CMS mode): Display and test record read correct.
- 18.  CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0813 Std Dev: 0.0006 Rel Std Dev: 0.68%

Maintenance/Repairs/Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Instrument is acceptable to be used in the field.

Charles E. Edr  
Field Inspector's Signature

3/23/17  
Date

Deb Shanaver  
Reviewed by

24 Mar 17  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = TOXL      8164.14.00 09/16  
03/22/2017      16:03

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.100  
Standard Lot #? 32114100A1  
Standard Cyl #? 27  
Standard Expiration? 12/05/2016  
Oper No? 666666

Flow Cal. Date: 09/07/2016  
Slope 698  
Intercept -642832

IR Calibration Date: 09/07/2016  
3um 9um

0th Coef(*100):	-10822	-23152
1st Coef(*100):	262468	134492
2nd Coef(*100):	3796	1436
H2O adj(ng/l*10K):	665	568

\*\*\*\*\* Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: *Internal*

*Print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                8164.14.00 09/16  
03/22/2017                      16:07

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            17316080A2  
Standard Cyl #?            26  
Standard Expiration?       08/05/2018  
Oper No?                    666666

Flow Cal. Date:            09/07/2016  
Slope                      698  
Intercept                  -642832

IR Calibration Date:       09/07/2016  
                                    3um                      9um

-----  
0th Coef(\*100):        -10822                -23152  
1st Coef(\*100):        262468                134492  
2nd Coef(\*100):        3796                    1436  
H2O adj(mg/l\*10k):    665                    568

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

EXTERNAL PRINT  
TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006688  
Location = TOXL                      8164.14.00 09/16  
03/22/2017                              16:08

Test	AC	Time
01 Room Air	0.000	16:09
02 *Subject Test	REF*	16:09
03 Room Air	0.000	16:10

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = R KEY REFUSAL  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                              Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *R KEY REFUSAL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/22/2017                            16:11

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *DIAGNOSTIC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/22/2017                            16:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:45
02 Std. Sol.	0.000	16:46
03 Room Air	0.000	16:47
04 Std. Sol.	0.000	16:47
05 Room Air	0.000	16:48
06 Std. Sol.	0.000	16:49
07 Room Air	0.000	16:49

08 Sim Temp = 34.0°C

Simul Ser No = MP3064  
Std Sol No = MILLIQ H2O  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

ZERO AC  
0.000 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = TOXL      8164.14.00 09/16  
03/22/2017      16:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:52
02 Std. Sol.	0.019	16:53
03 Room Air	0.000	16:53
04 Std. Sol.	0.019	16:54
05 Room Air	0.000	16:55
06 Std. Sol.	0.019	16:55
07 Room Air	0.000	16:56

08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 201507F

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*

*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = TOXL      8164.14.00 09/16  
03/23/2017      09:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:40
02 Std. Sol.	0.391	09:41
03 Room Air	0.000	09:41
04 Std. Sol.	0.392	09:42
05 Room Air	0.000	09:42
06 Std. Sol.	0.391	09:43
07 Room Air	0.000	09:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3063

Std Sol No = 16102

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC

0.400 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = TOXL      8164.14.00 09/16  
03/23/2017      10:30

Test	AC	Time
01 Room Air	0.000	10:31
02 *Subject Test	INT*	10:31
03 Room Air	0.000	10:32

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                8164.14.00 09/16  
03/23/2017                        10:33

Test	AC	Time
01 Room Air	0.000	10:34
02 *Subject Test	INT*	10:34
03 Room Air	0.000	10:35

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: 0.05% ACETONE



Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/23/2017                            10:35

Test	AC	Time
01 Room Air	0.000	10:36
02 *Subject Test	INT*	10:36
03 Room Air	0.000	10:37

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*

*0.10% METHANOL*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/23/2017                            10:38

Test	AC	Time
01 Room Air	0.000	10:39
02 *Subject Test	INT*	10:39
03 Room Air	0.000	10:40

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*

*0.10% ISOPROPNOL*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006688  
Location = TOXL                      8164.14.00 09/16  
03/23/2017    10:41

Test	AC	Time
01 Room Air	AMB*	10:42
02 Room Air	0.000	10:42

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                              Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*AMBIENT FAIL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/23/2017                            10:43

Test	AC	Time
01 Room Air	0.000	10:44
02 Invalid Sample	X.XXX	10:44
03 Room Air	0.000	10:45

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INVALID SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

INVALID  
SAMPLE

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/23/2017                            10:45

Test	AC	Time
01 Room Air	0.000	10:46
02 *Subject Test	ABT*	10:46
03 Room Air	0.000	10:47

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: SEQUENCE ABORTED


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                8164.14.00 09/16  
03/23/2017                        10:47

Test	AC	Time
01 Room Air	0.000	10:47
02 Subject Test 1	0.019	10:48
03 Room Air	0.000	10:48
04 Reported AC	0.019	10:48

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = ABA MODE  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                          Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *ABA MODE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-006688  
Location = TOXL                      8164.14.00 09/16  
03/23/2017                      10:49

Test	AC	Time
01 Diagnostic	OK	10:50
02 Room Air	0.000	10:51
03 Subject Test 1	0.010	10:51
04 Room Air	0.000	10:53
05 Std. Gas	0.090*	10:54
06 Room Air	0.000	10:55

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                      Cit = CAL CHK OUT TOL  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                      Oper No. = 666666

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
OUT OF TOLERANCE

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                 8164.14.00 09/16  
03/23/2017                        10:56

Test	AC	Time
01 Diagnostic	OK	10:57
02 Room Air	RFI*	10:57
03 Room Air	0.000	10:58

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI DETECTED  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI DETECTED

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/23/2017                            10:59

Test	AC	Time
01 Diagnostic	OK	11:00
02 Room Air	0.000	11:00
03 *Subject Test	0.000*	11:03
04 Room Air	0.000	11:06
05 Std. Gas	0.081	11:07
06 Room Air	0.000	11:08
07 *Subject Test	0.000*	11:11
08 Room Air	0.000	11:12


No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DEFICIENT SAMP  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

DEFICIENT  
SAMPLES

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006688  
Location = TOXL                      8164.14.00 09/16  
03/23/2017                              11:13

Test	AC	Time
01 Diagnostic	OK	11:13
02 Room Air	0.000	11:14
03 Subject Test 1	0.000	11:14
04 Room Air	0.000	11:16
05 Std. Gas	0.081	11:17
06 Room Air	0.000	11:19
07 Reference	IPS*	11:19
08 Room Air	0.000	11:19

\*Improper Sample

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = IMPROPER SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                              Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*Improper  
Sample*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                8164.14.00 09/16  
03/23/2017                        12:16

Test	AC	Time
01 Diagnostic	OK	12:17
02 Room Air	0.000	12:17
03 Subject Test 1	0.146*	12:18
04 Room Air	0.000	12:20
05 Std. Gas	0.082	12:21
06 Room Air	0.000	12:23
07 Subject Test 2	0.099*	12:23
08 Room Air	0.000	12:24

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *DIFFERENCE*  
*TOO GREAT*  
Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006688  
Location = TOXL      8164.14.00 09/16  
03/23/2017      12:26


Test	AC	Time
01 Diagnostic	OK	12:27
02 Room Air	0.000	12:28
03 Subject Test 1	0.143	12:28
04 Room Air	0.000	12:30
05 Std. Gas	0.081	12:31
06 Room Air	0.000	12:33
07 Subject Test 2	0.143	12:33
08 Room Air	0.000	12:34
09 Reported AC	0.143	12:28

Difference OK  
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = CMS MODE  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08      Oper No. = 666666

20 minute waiting period ascertained? Y  
Clean Mouthpiece used and disposed? Y  
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the  
instructions displayed by the Intoxilyzer  
in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

CMS MODE



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006688  
Location = TOXL                    8164.14.00 09/16  
03/23/2017                            11:20

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.081	11:21
03 Room Air	0.000	11:22
04 Std. Gas	0.081	11:22
05 Room Air	0.000	11:22
06 Std. Gas	0.081	11:23
07 Room Air	0.000	11:23
08 Std. Gas	0.081	11:24
09 Room Air	0.000	11:24
10 Std. Gas	0.081	11:25
11 Room Air	0.000	11:25
12 Std. Gas	0.081	11:25
13 Room Air	0.000	11:26
14 Std. Gas	0.081	11:26
15 Room Air	0.000	11:27
16 Std. Gas	0.081	11:27
17 Room Air	0.000	11:28
18 Std. Gas	0.082	11:28
19 Room Air	0.000	11:29
20 Std. Gas	0.081	11:29
21 Room Air	0.000	11:29
22 Std. Gas	0.081	11:30
23 Room Air	0.000	11:30
24 Std. Gas	0.081	11:31
25 Room Air	0.000	11:31
26 Std. Gas	0.082	11:31
27 Room Air	0.000	11:32
28 Std. Gas	0.082	11:32
29 Room Air	0.000	11:33
30 Std. Gas	0.082	11:33
31 Room Air	0.000	11:34
32 Std. Gas	0.082	11:34
33 Room Air	0.000	11:35
34 Std. Gas	0.081	11:35
35 Room Air	0.000	11:35
36 Std. Gas	0.081	11:36
37 Room Air	0.000	11:36
38 Std. Gas	0.080	11:37
39 Room Air	0.000	11:37
40 Std. Gas	0.082	11:38
41 Room Air	0.000	11:38
42 Cal Check Stats:		
43 Average	0.0813	
44 Std Dev	0.0006	
45 Rel Std Dev	0.68%	



Operator Signature  
CHARLES EDER

Remarks: STABILITY TEST

0.080 AC

Form 106-I8000

Lot: 17316080A2

CYL: 026

Exp: 08-05-2018