

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006667

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN MP3003 Lot No. 201610A Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN MP3063 Lot No. 16102 Rep AC 0.400 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5135 Lot No. FCS4 Solution 0.10% AC + 0.05% ACETONE ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 17316080A2 Cyl No. 026 Exp. Date 8/5/2018

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.082</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 377 psi Reg 400 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

4.6.2017  
Date

Charles E. Ede  
Field Inspector's Signature

11 Apr 17  
Date

Deb Shanaver  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006667  
Location = TOXL                    8164.14.00 09/16  
04/06/2017                            11:06

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            17316080A2  
Standard Cyl #?            26  
Standard Expiration?       08/05/2018  
Oper No?                    666666

Flow Cal. Date:            08/15/2016  
Slope                        699  
Intercept                   -645669

IR Calibration Date:       08/15/2016

	3um	9um
0th Coef(*100):	-23625	-25093
1st Coef(*100):	269805	129386
2nd Coef(*100):	3177	1385
H2O adj(mg/l*10k):	621	546

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: *PRINT TEST*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = TOXL      8164.14.00 09/16  
04/06/2017      11:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Sol.	0.017	11:11
03 Room Air	0.000	11:12
04 Std. Sol.	0.018	11:13
05 Room Air	0.000	11:13
06 Std. Sol.	0.018	11:14
07 Room Air	0.000	11:14

08 Sim Temp = 34.0°C

Simul Ser No = MP3003  
Std Sol No = 201610A  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*  
*0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = TOXL      8164.14.00 09/16  
04/06/2017      11:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:17
02 Std. Sol.	0.398	11:17
03 Room Air	0.000	11:18
04 Std. Sol.	0.400	11:19
05 Room Air	0.000	11:19
06 Std. Sol.	0.399	11:20
07 Room Air	0.000	11:20

08 Sim Temp = 34.0°C

Simul Ser No = MP3063

Std Sol No = 16102

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = TOXL      8164.14.00 09/16  
04/06/2017      12:12

Test	AC	Time
01 Room Air	0.000	12:13
02 *Subject Test	INT*	12:14
03 Room Air	0.000	12:14

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 26  
Expiration Date = 08/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.10 AC + 0.05% ACETONE*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = TOXL      8164.14.00 09/16  
04/06/2017      12:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:18
02 Std. Gas	0.081	12:18
03 Room Air	0.000	12:19
04 Std. Gas	0.082	12:19
05 Room Air	0.000	12:20
06 Std. Gas	0.081	12:20
07 Room Air	0.000	12:21

Lot No = 17316080A2  
Cyl No = 26  
Exp Date = 08/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = TOXL      8164.14.00 09/16  
04/06/2017      12:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:22
02 Std. Gas	0.081	12:22
03 Room Air	0.000	12:23
04 Std. Gas	0.081	12:23
05 Room Air	0.000	12:23
06 Std. Gas	0.082	12:24
07 Room Air	0.000	12:24

Lot No = 17316080A2  
Cyl No = 26  
Exp Date = 08/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006667  
Location = TOXL      8164.14.00 09/16  
04/06/2017      12:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:26
02 Std. Gas	0.081	12:26
03 Room Air	0.000	12:26
04 Std. Gas	0.081	12:27
05 Room Air	0.000	12:27
06 Std. Gas	0.081	12:28
07 Room Air	0.000	12:28

Lot No = 17316080A2  
Cyl No = 26  
Exp Date = 08/05/2018  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000