

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006689

Location: TOXL

- Check When Done
- A. Is the breath tube heated? X
- B. Is the diagnostic check complete? X
- C. Is the time, date, and year correct (re-set if necessary)? X
- D. Print test. (Attach test record.) X
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3846 Lot No. 160200 Rep AC 0.020 X
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR7111 Lot No. 16060 Rep AC 0.300 X
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN G10329 Lot No. 1CS4 Solution 0.10 AC + 0.05% X
- H. RFI check. Run CMS mode. Key radio during CMS test. Acetone X
(Attach test record.)
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 17316080A2 Cyl No. 27 Exp. Date 5 Aug 18

Test 1 0.082 Test 4 0.082 Test 7 0.083

Test 2 0.082 Test 5 0.082 Test 8 0.082

Test 3 0.082 Test 6 0.082 Test 9 0.083

Average 0.082

J. Tank pressure, Level 3 Func D Sub G

933 950
P: 950 psi Reg 980 psi

K. Remarks/Maintenance record: Annual Inspection.

17 Apr 17
Date

Roberta Mungo-Nimms
Field Inspector's Signature

17 Apr 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 09:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Sol.	0.021	09:46
03 Room Air	0.000	09:46
04 Std. Sol.	0.021	09:47
05 Room Air	0.000	09:48
06 Std. Sol.	0.021	09:48
07 Room Air	0.000	09:49

08 Sim Temp = 34.0°C

Simul Ser No = DR3846
Std Sol No = 16200
County = 08 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 09:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:51
02 Std. Sol.	0.300	09:52
03 Room Air	0.000	09:52
04 Std. Sol.	0.301	09:53
05 Room Air	0.000	09:53
06 Std. Sol.	0.302	09:54
07 Room Air	0.000	09:55

08 Sim Temp = 34.0°C

Simul Ser No = DR7111
Std Sol No = 16060
County = 08 Oper No. = 888888

Roberta Grieger-Nimmo
Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: High AC
 0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 10:03

Test	AC	Time
01 Room Air	0.000	10:04
02 *Subject Test	INT*	10:05
03 Room Air	0.000	10:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 27
Expiration Date = 08/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Interferent Check
0.10 AC + 0.05% Acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 10:06

Test	AC	Time
01 Diagnostic	OK	10:07
02 Room Air	RFI*	10:07
03 Room Air	0.000	10:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 27
Expiration Date = 08/05/2018
County = 08 Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 10:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Gas	0.082	10:32
03 Room Air	0.000	10:33
04 Std. Gas	0.082	10:33
05 Room Air	0.000	10:34
06 Std. Gas	0.082	10:34
07 Room Air	0.000	10:35

Lot No = 17316080A2
Cyl No = 27
Exp Date = 08/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC
Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:35
02 Std. Gas	0.082	10:36
03 Room Air	0.000	10:36
04 Std. Gas	0.082	10:37
05 Room Air	0.000	10:37
06 Std. Gas	0.082	10:37
07 Room Air	0.000	10:38

Lot No = 17316080A2
Cyl No = 27
Exp Date = 08/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006689
Location = TOXL 8164.14.00 09/16
04/17/2017 10:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:39
02 Std. Gas	0.083	10:39
03 Room Air	0.000	10:40
04 Std. Gas	0.082	10:40
05 Room Air	0.000	10:41
06 Std. Gas	0.083	10:41
07 Room Air	0.000	10:42

Lot No = 17316080A2
Cyl No = 27
Exp Date = 08/05/2018
County = 08 Oper No. = 888888

Roberta Grieger-Nimmo

Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*

0.080 AC

Form 106-I8000