



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (10-2016)

Serial Number <i>80-006679</i>	Instrument Location <i>VALL</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Dawn Russell</i>	Date <i>5/8/17</i>
--	-----------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified by <i>Deb Shanaver</i>	Certified Date <i>10 May 17</i>
--	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/08/2017 16:58

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01316080A2
Standard Cyl #? 31
Standard Expiration? 02/05/2018
Oper No? 100012

Flow Cal. Date: 09/06/2016
Slope 701
Intercept -682926

IR Calibration Date: 09/06/2016
 3um 9um

0th Coef(*100):	-18983	-23165
1st Coef(*100):	280986	134481
2nd Coef(*100):	2443	1290
H2O adj(mg/l*10k):	569	553

***** Printer Test End *****



Operator Signature
DANA RUSTEBAKKE

Remarks:

Print test


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/08/2017 16:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:59
02 Std. Gas	0.079	16:59
03 Room Air	0.000	17:00
04 Std. Gas	0.079	17:00
05 Room Air	0.000	17:01
06 Std. Gas	0.078	17:01
07 Room Air	0.000	17:01

Lot No = 01316080A2
Cyl No = 31
Exp Date = 02/05/2018
County = 02 Oper No. = 100012



Operator Signature
DANA RUSTEBAKKE

Remarks:

ACA Test

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
05/08/2017 17:03

Test	AC	Time
01 Room Air	0.000	17:04
02 *Subject Test	RFI*	17:04
03 Room Air	0.000	17:05

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/01/1980
Sub Sex = Male Weight = 150
Test = DUI Cit = 123456
Dr. Lic. = ND/TEST801234
Lot No = 01316080A2
Cyl No = 31
Expiration Date = 02/05/2018
County = 02 Oper No. = 100012



Operator Signature
DANA RUSTEBAKKE

Remarks:

RFI TEST

Form 106-I8000