

INTOXILYZER® 8000 INSPECTION Foster Co. S.O.

Serial No.: 80-006512

Location: CARRINGTON P.D (CARR)

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5113</u> Lot No. <u>201507F</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3378</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>IC54</u> Solution <u>0.10AC + 0.05%</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. 19615080A5 Cyl No. 28 Exp. Date 9/5/2017

Test 1 0.079 Test 4 0.079 Test 7 0.080

Test 2 0.079 Test 5 0.080 Test 8 0.080

Test 3 0.079 Test 6 0.079 Test 9 0.079

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 1002 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/12/2017
Date

Charles E. Eds
Field Inspector's Signature

12 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 28
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 08/20/2015
Slope 680
Intercept -660856

IR Calibration Date: 08/21/2015
 3um 9um

	3um	9um
0th Coef(*100):	-23051	-19261
1st Coef(*100):	266836	137135
2nd Coef(*100):	3974	1466
H2O adj(mg/l*10k):	785	487

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:35
02 Std. Sol.	0.020	11:36
03 Room Air	0.000	11:36
04 Std. Sol.	0.019	11:37
05 Room Air	0.000	11:38
06 Std. Sol.	0.020	11:38
07 Room Air	0.000	11:39


08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 16

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LOW AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Sol.	0.305	11:41
03 Room Air	0.000	11:42
04 Std. Sol.	0.306	11:43
05 Room Air	0.000	11:43
06 Std. Sol.	0.306	11:44
07 Room Air	0.000	11:45

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 16

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:46

Test	AC	Time
01 Room Air	0.000	11:47
02 *Subject Test	INT*	11:48
03 Room Air	0.000	11:48

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 28
Expiration Date = 09/05/2017
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:49

Test	AC	Time
01 Diagnostic	OK	11:49
02 Room Air	0.000	11:50
03 *Subject Test	RFI*	11:50
04 Room Air	0.000	11:50

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 28
Expiration Date = 09/05/2017
County = 16 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:52
02 Std. Gas	0.079	11:52
03 Room Air	0.000	11:53
04 Std. Gas	0.079	11:53
05 Room Air	0.000	11:54
06 Std. Gas	0.079	11:54
07 Room Air	0.000	11:54

Lot No = 19615080A5
Cyl No = 28
Exp Date = 09/05/2017
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Gas	0.079	11:55
03 Room Air	0.000	11:56
04 Std. Gas	0.080	11:56
05 Room Air	0.000	11:57
06 Std. Gas	0.079	11:57
07 Room Air	0.000	11:58

Lot No = 19615080A5
Cyl No = 28
Exp Date = 09/05/2017
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006512
Location = CARR 8164.14.00 09/16
05/12/2017 11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Gas	0.080	11:59
03 Room Air	0.000	12:00
04 Std. Gas	0.080	12:00
05 Room Air	0.000	12:00
06 Std. Gas	0.079	12:01
07 Room Air	0.000	12:01

Lot No = 19615080A5
Cyl No = 28
Exp Date = 09/05/2017
County = 16 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000