



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number <i>80-006686</i>	Instrument Location <i>HP office - Walsh Co SO</i> <i>Deb Shanaver</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *None*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

<i>Ben Winger</i> #1442	<i>9/12/2017</i>
Field Inspector Signature	Date
<i>Ben Winger</i> #1442	<i>5-17-2017</i>
Reviewed By (Crime Laboratory Use Only)	Date
<i>Deb Shanaver</i>	<i>26 May 17</i>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Robata Luque-Tumme* Certified Date: *19 Sep 17*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/17/2017 18:17

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? ~~17316080A2~~ 01316080A2
Standard Cyl #? ~~28~~ 27
Standard Expiration? 08/05/2018
Oper No? 130890

BJW-9-12-17
27 - BJW - 9-12-17

Flow Cal. Date: 09/07/2016
Slope 770
Intercept -749410

IR Calibration Date: 09/06/2016

	3um	9um
0th Coef(*100):	-14519	-24061
1st Coef(*100):	252211	131592
2nd Coef(*100):	3527	1490
H2O adj(mg/l*10k):	522	594

***** Printer Test End *****

Bradley Wingarter #1442 - 9/12/2017
Bradley Wingarter #1442

Operator Signature
BRADLEY WINGERTER

Remarks: *Noted Improper Gas cylinder numbers*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/17/2017 18:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:18
02 Std. Gas	0.080	18:18
03 Room Air	0.000	18:19
04 Std. Gas	0.079	18:19
05 Room Air	0.000	18:20
06 Std. Gas	0.079	18:20
07 Room Air	0.000	18:21

Lot No = ~~17316080A2~~ 01316080A2 - BSW - 9/12/2017
Cyl No = ~~28~~ 27 - BSW - 9/12/2017
Exp Date = 08/05/2018
County = 50 Oper No. = 130890

Bradley Wingarter #1442 - 9/12/2017
Bradley Wingarter #1442

Operator Signature
BRADLEY WINGERTER

Remarks: *Noted Improper Gas Cylinder numbers*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006686
Location = GRAF 8164.14.00 09/16
05/17/2017 18:23

Test	AC	Time
01 Room Air	RFI*	18:25
02 Room Air	0.000	18:26

*Invalid Test
Inhibited - RFI

Sub Name = DOE, JANE NA
Sub DOB = 12/12/1972
Sub Sex = Female Weight = 135
Test = DUI Cit = 1234567
Dr. Lic. = ND/DOE724534
Lot No = ~~47346080A2~~ 01316080A2 - BJW - 9/12/2017
Cyl No = ~~28~~ 27 - BJW - 9/12/2017
Expiration Date = 08/05/2018
County = 50 Oper No. = 130890

Brad Wingarter #1442 9/12/2017

Brad Wingarter #1442

Operator Signature
BRADLEY WINGERTER

Remarks: *Noted Improper Gas Cylinder Numbers*