

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006509

Location: TOX L

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3003 Lot No. 16200 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3063 Lot No. 16102 Rep AC 0.400 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. JCS4 Solution 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 17316080A2 Cyl No. 28 Exp. Date 8/5/2018

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.080</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 374 psi Reg 325 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/26/2017
Date

30 May 17
Date

Charles E. Edler
Field Inspector's Signature

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
05/26/2017 16:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:02
02 Std. Sol.	0.017	16:03
03 Room Air	0.000	16:03
04 Std. Sol.	0.018	16:04
05 Room Air	0.000	16:04
06 Std. Sol.	0.018	16:05
07 Room Air	0.000	16:06

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 16200
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
05/26/2017 16:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:07
02 Std. Sol.	0.399	16:08
03 Room Air	0.000	16:09
04 Std. Sol.	0.400	16:09
05 Room Air	0.000	16:10
06 Std. Sol.	0.399	16:11
07 Room Air	0.000	16:11

08 Sim Temp = 34.0°C

Simul Ser No = MP3063
Std Sol No = 16102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
05/26/2017 16:13

Test	AC	Time
01 Room Air	0.000	16:14
02 *Subject Test	INT*	16:14
03 Room Air	0.000	16:15

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 28
Expiration Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent
0.10AC + 0.05% ACETONE


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
05/26/2017 16:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:18
02 Std. Gas	0.079	16:18
03 Room Air	0.000	16:19
04 Std. Gas	0.080	16:19
05 Room Air	0.000	16:20
06 Std. Gas	0.079	16:20
07 Room Air	0.000	16:21

Lot No = 17316080A2
Cyl No = 28
Exp Date = 08/05/2018
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
05/26/2017 16:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:21
02 Std. Gas	0.079	16:22
03 Room Air	0.000	16:22
04 Std. Gas	0.079	16:23
05 Room Air	0.000	16:23
06 Std. Gas	0.079	16:24
07 Room Air	0.000	16:24

Lot No = 17316080A2
Cyl No = 28
Exp Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006509
Location = TOXL 8164.14.00 09/16
05/26/2017 16:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Gas	0.079	16:25
03 Room Air	0.000	16:26
04 Std. Gas	0.080	16:26
05 Room Air	0.000	16:27
06 Std. Gas	0.080	16:27
07 Room Air	0.000	16:28

Lot No = 17316080A2
Cyl No = 28
Exp Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC