



SERVICE WORK ORDER

DATE	3-21-17
RCVD VIA	FEDEX

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

340992

BILL TO:		
ATTN:		
CITY STATE ZIP		
CUSTOMER NO. 585013		
PHONE		

SHIP TO: ND Crime Lab		
ATTN:		
2041 E. Main Ave.		
CITY	STATE	ZIP
Bismarck	ND	58501
PHONE 701-328-6159		
FAX 701-328-6185		

MODEL # 8000	SERIAL # 00-006672	BILL CODE 3
--------------	--------------------	-------------

EXTRA PARTS RCVD: BT, H, gas box, Battery

DESCRIPTION OF PROBLEM: Frequent voltage/current fails

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION

BATTERY- EXT LOW V; INT 3.065V
 OK ✓

VERIFIED PROBLEM? yes

WORK PERFORMED:
 Replaced 12V BTY with new cramps

Miscellaneous
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:
 BT

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	013010	12V BTY
4	320384	cramps BTY

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
DU	DDP	DDP	DDP
½ HOURS	HOURS	HOURS	HOURS
6	0.5	1.0	1.0
DATE	DATE	DATE	DATE
June 29, 2017	July 5, 17	July 5, 17	July 5, 17

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS