

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006672

Agency being assigned instrument: OFFICE OF ATTORNEY GENERAL -

CRIME LAB

Date 7/12/17

Transferred from:

Name (print) CMI, INC

Signature _____

Agency CMI, INC

Transferred to:

Name (print) CHARLES E. EDER

Signature Charles E. Eder

Agency O. AG. - CRIME LAB

Reason for transfer/remarks:

- RETURNED FROM CMI AFTER WARRANTY REPAIR
- VOLTAGE FAILS INTERMITTENT ON AC POWER, VOLTAGE FAILS EVERYTIME ON BATTERY ALONE
- CMI REPLACED 12V BATTERY WITH NEW CRIMPER

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____