

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006674

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 201603H Rep AC 0.010 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3063 Lot No. 201603A Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR3451 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 03417080A3 Cyl No. 027 Exp. Date 3/5/2019

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.080</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 965 psi Reg 950 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

7/20/17
Date

Chadwick E. Lee
Field Inspector's Signature

20 July 17
Date

Roberta Henger-Timm
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 15:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 27
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 06/28/2017
Slope 707
Intercept -708478

IR Calibration Date: 06/28/2017

	3um	9um
0th Coef(*100):	-3915	-18046
1st Coef(*100):	256520	127506
2nd Coef(*100):	3843	1644
H2O adj(mg/l*10k):	428	568

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 15:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:47
02 Std. Sol.	0.010	15:48
03 Room Air	0.000	15:48
04 Std. Sol.	0.010	15:49
05 Room Air	0.000	15:49
06 Std. Sol.	0.010	15:50
07 Room Air	0.000	15:50

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 201603H
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.010 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 15:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:53
02 Std. Sol.	0.299	15:54
03 Room Air	0.000	15:54
04 Std. Sol.	0.300	15:55
05 Room Air	0.000	15:56
06 Std. Sol.	0.299	15:56
07 Room Air	0.000	15:57

08 Sim Temp = 34.0°C

Simul Ser No = MP3063
Std Sol No = 201603A
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 15:38

Test	AC	Time
01 Room Air	0.000	15:39
02 *Subject Test	INT*	15:39
03 Room Air	0.000	15:40

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 15:35

Test	AC	Time
01 Diagnostic	OK	15:36
02 Room Air	0.000	15:36
03 *Subject Test	RFI*	15:36
04 Room Air	0.000	15:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 16:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:03
02 Std. Gas	0.081	16:03
03 Room Air	0.000	16:04
04 Std. Gas	0.081	16:04
05 Room Air	0.000	16:05
06 Std. Gas	0.081	16:05
07 Room Air	0.000	16:05

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 16:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:06
02 Std. Gas	0.081	16:07
03 Room Air	0.000	16:07
04 Std. Gas	0.081	16:08
05 Room Air	0.000	16:08
06 Std. Gas	0.081	16:09
07 Room Air	0.000	16:09

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = TOXL 8164.14.00 09/16
07/20/2017 16:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:10
02 Std. Gas	0.080	16:10
03 Room Air	0.000	16:11
04 Std. Gas	0.081	16:11
05 Room Air	0.000	16:12
06 Std. Gas	0.081	16:12
07 Room Air	0.000	16:13

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000