



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (10-2016)

Serial Number 80-006674	Instrument Location B.M.D.C.
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P). ✓
 - B. ACA Test (Level 1, Function C). ✓
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test). ✓
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

RFI Checked box on 23 Aug 17 as tests were completed on 8/11/17

Field Inspector Signature Rebecca Kopp	Date 8-11-17
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified by Roberta Huger-Nimmro	Certified Date 23 Aug 17
--	------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = BMCD 8164.14.00 09/16
08/11/2017 09:58

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 20016080A2
Standard Cyl #? 45
Standard Expiration? 09/05/2018
Oper No? 041005

Flow Cal. Date: 06/28/2017
Slope 707
Intercept -708478

IR Calibration Date: 06/28/2017
 3um 9um

	3um	9um
0th Coef(*100):	-3915	-18046
1st Coef(*100):	256520	127506
2nd Coef(*100):	3843	1644
H2O adj(mg/l*10k):	428	568

***** Printer Test End *****



Operator Signature
REBECCA KOPP

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = BMCD 8164.14.00 09/16
08/11/2017 10:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Gas	0.080	10:01
03 Room Air	0.000	10:01
04 Std. Gas	0.080	10:02
05 Room Air	0.000	10:02
06 Std. Gas	0.080	10:03
07 Room Air	0.000	10:03

Lot No = 20016080A2
Cyl No = 45
Exp Date = 09/05/2018
County = 08 Oper No. = 041005



Operator Signature
REBECCA KOPP

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006674
Location = BMCD 8164.14.00 09/16
08/11/2017 10:04

Test	AC	Time
01 Room Air	RFI*	10:05
02 Room Air	0.000	10:06

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST NA
Sub DOB = 01/01/1999
Sub Sex = Female Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 20016080A2
Cyl No = 45
Expiration Date = 09/05/2018
County = 08 Oper No. = 041005



Operator Signature
REBECCA KOPP

Remarks: