



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (10-2016)

Serial Number <i>80-006672</i>	Instrument Location <i>NTPD</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *N/A*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>09/07/2017</i>
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified by <i>Roberta Senger-Nimmo</i>	Certified Date <i>09/07/2017</i>
--	-------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
09/02/2017 18:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 30
Standard Expiration? 03/05/2019
Oper No? 130826

Flow Cal. Date: 07/05/2017
Slope 705
Intercept -687158

IR Calibration Date: 07/05/2017
 3um 9um

0th Coef(*100): -22113 -17955
1st Coef(*100): 271833 130091
2nd Coef(*100): 2129 1363
H2O adj(mg/l*10k): 691 568

***** Printer Test End *****



Operator Signature
JACQUELYN HALONEN

Remarks:

Print Test


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
09/02/2017 18:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:48
02 Std. Gas	0.080	18:48
03 Room Air	0.000	18:49
04 Std. Gas	0.080	18:49
05 Room Air	0.000	18:49
06 Std. Gas	0.080	18:50
07 Room Air	0.000	18:50

Lot No = 03417080A3
Cyl No = 30
Exp Date = 03/05/2019
County = 31 Oper No. = 130826


Operator Signature
JACQUELYN HALONEN

Remarks:

ACA TEST


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
09/02/2017 18:52

Test	AC	Time
01 Room Air	0.000	18:56
02 *Subject Test	RFI*	18:56
03 Room Air	0.000	18:56

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 10/30/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = TEST922017
Dr. Lic. = ND/TEST20170902
Lot No = 03417080A3
Cyl No = 30
Expiration Date = 03/05/2019
County = 31 Oper No. = 130826



Operator Signature
JACQUELYN HALONEN

Remarks:

RFI Test

Form 106-I8000