



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (10-2016)

Serial Number <i>80006672</i>	Instrument Location <i>NTPD</i>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): *NA*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>12/26/2017</i>
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified by <i>Roberta Sawyer-Munim</i>	Certified Date <i>3 Jan 18</i>
--	-----------------------------------

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006672  
Location = NTPD      8164.14.00 09/16  
12/26/2017      17:38

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      03417080A3  
Standard Cyl #?      30  
Standard Expiration?      03/05/2019  
Oper No?      130826  
  
Flow Cal. Date:      07/05/2017  
Slope      705  
Intercept      -687158

IR Calibration Date:      07/05/2017

	3um	9um
0th Coef(*100):	-22113	-17955
1st Coef(*100):	271833	130091
2nd Coef(*100):	2129	1363
H2O adj(mg/l*10k):	691	568

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JACQUELYN HALONEN

Remarks:

Print test: following move-install


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006672  
Location = NTPD      8164.14.00 09/16  
12/26/2017      17:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:43
02 Std. Gas	0.083	17:44
03 Room Air	0.000	17:44
04 Std. Gas	0.083	17:45
05 Room Air	0.000	17:45
06 Std. Gas	0.083	17:45
07 Room Air	0.000	17:46

Lot No = 03417080A3  
Cyl No = 30  
Exp Date = 03/05/2019  
County = 31      Oper No. = 130826

  
Operator Signature  
JACQUELYN HALONEN

Remarks:

ACA Test: Install after Intoxilyzer was moved

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006672  
Location = NTPD      8164.14.00 09/16  
12/26/2017      17:55

Test	AC	Time
01 Room Air	0.000	17:56
02 *Subject Test	RFI*	17:57
03 Room Air	0.000	17:57

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, TEST TEST  
Sub DOB = 12/26/1990  
Sub Sex = Male      Weight = 150  
Test = DUI      Cit = 20161226  
Dr. Lic. = ND/TEST1234  
Lot No = 03417080A3  
Cyl No = 30  
Expiration Date = 03/05/2019  
County = 31      Oper No. = 130826



Operator Signature  
JACQUELYN HALONEN

Remarks:

RFI Test: In stall following Intoxilyzer move