



**ETHANOL GAS STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Bradley Wingeter</i>		
Location <i>Walsh County</i>		Intoxilyzer® Serial Number <i>80-006686</i>
Gas Lot Number <i>19817080A4</i>	Gas Cylinder Number <i>22</i>	Gas Expiration Date <i>9-5-2019</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.083 AC
  - B. 0.083 AC
  - C. 0.083 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Bradley Wingeter</i>	Date <i>2-1-2018</i>
Reviewed By (Crime Laboratory Use Only) <i>Robert Muegge Nimmo</i>	Date <i>2-5-2018</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006686  
Location = GRAF      8164.14.00 09/16  
02/01/2018      13:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:04
02 Std. Gas	0.083	13:04
03 Room Air	0.000	13:05
04 Std. Gas	0.083	13:05
05 Room Air	0.000	13:05
06 Std. Gas	0.083	13:06
07 Room Air	0.000	13:06


Lot No = 19817080A4

Cyl No = 22

Exp Date = 09/05/2019

County = 50

Oper No. = 130890



Operator Signature

BRADLEY WINGERTER

Remarks: Calibration Check

Form 106-I8000