



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Dana Kustebekke</i>		
Location <i>VALL</i>	Intoxilyzer® Serial Number <i>80-006679</i>	
Gas Lot Number <i>19817080A4</i>	Gas Cylinder Number <i>6</i>	Gas Expiration Date <i>9/5/2019</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.081* AC
 - B. *0.082* AC
 - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Dana Kustell</i>	Date <i>2/5/2018</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Juergens-Nemmo</i>	Date <i>2/5/2018</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006679
Location = VALL 8164.14.00 09/16
02/05/2018 14:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.081	14:40
03 Room Air	0.000	14:41
04 Std. Gas	0.082	14:41
05 Room Air	0.000	14:42
06 Std. Gas	0.082	14:42
07 Room Air	0.000	14:43

Lot No = 19817080A4
Cyl No = 6
Exp Date = 09/05/2019
County = 02 Oper No. = 100012



Operator Signature
DANA RUSTEBAKKE

Remarks:

ACA Test
New Gas Cylinder
Form 106-I8000