



### ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Jeff Fleck</i>		
Location <i>La Moure Sheriff's Dept</i>	Intoxilyzer® Serial Number <i>80-006680</i>	
Gas Lot Number <i>19817080A4</i>	Gas Cylinder Number <i>7</i>	Gas Expiration Date <i>9/5/2019</i>

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C).

*ACA test  
was performed  
02/06/2018.  
RSN.*

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. *0.082* AC

B. *0.081* AC

C. *0.080* AC

3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature <i>Jeff Fleck</i>	Date <i>02/06/2018</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Hunger-Numms</i>	Date <i>02/07/2018</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006680  
Location = LAMO                8164.14.00 09/16  
02/06/2018                      11:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:14
02 Std. Gas	0.082	11:15
03 Room Air	0.000	11:15
04 Std. Gas	0.081	11:16
05 Room Air	0.000	11:16
06 Std. Gas	0.080	11:17
07 Room Air	0.000	11:17

Lot No = 19817080A4  
Cyl No = 7  
Exp Date = 09/05/2019  
County = 23                      Oper No. = 010605

  
Operator Signature  
JEFFREY D. FLECK

Remarks:

LaMoure County Sheriff's Department  
PO Box 97  
LaMoure, ND 58458  
Phone: (701) 883 - 5720, ext. 1  
Fax: (701) 883 - 5833

# facsimile transmittal

To: \_\_\_\_\_ Fax: \_\_\_\_\_


From: \_\_\_\_\_ Date: \_\_\_\_\_

Re: \_\_\_\_\_ Pages: \_\_\_\_\_

CC: \_\_\_\_\_

- Urgent
- For Review
- Please Comment
- Please Reply
- Please Recycle

Note: If any of these fax copies are illegible, or you do not receive the same number of pages as stated above, please contact us immediately at (701) 883-5720, ext. 1



*Attn  
Breath  
Analysis*

