



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Elli, Anthony C</i>		
Location <i>WCPD</i>	Intoxilyzer® Serial Number <i>80-006673</i>	
Gas Lot Number <i>19817080A4</i>	Gas Cylinder Number <i>020</i>	Gas Expiration Date <i>09-05-2019</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.080* AC
B. *0.080* AC
C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Anthony C Elli</i>	Date <i>3-23-18</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Yungue-Timmo</i>	Date <i>3-26-18</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = WCPD 8164.14.00 09/16
03/23/2018 15:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:55
02 Std. Gas	0.080	15:55
03 Room Air	0.000	15:56
04 Std. Gas	0.080	15:56
05 Room Air	0.000	15:56
06 Std. Gas	0.080	15:57
07 Room Air	0.000	15:57

Lot No = 19817080A4
Cyl No = 20
Exp Date = 09/05/2019
County = 27 Oper No. = 131814



Operator Signature
ANTHONY ELL

Remarks: