

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006675

Location: BSPD  
(BISMARCK P.D.)

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 234 psi Regulator: 225 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10/11/19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR345 Lot #: ICS #5 Exp. Date: 5.15.19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 17316080A2 Cyl #: 25 Exp. Date: 8/5/2018

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.080</u>	Test 6	<u>0.081</u>	Test 9	<u>0.080</u>
Average <u>0.081</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr  
Breath Analyst Signature

4.20.2018  
Date

Robert Guegu-Timmer  
Reviewed by

4.20.2018  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006675  
Location = BSPD                8164.14.00 09/16  
04/20/2018                      12:56

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            17316080A2  
Standard Cyl #?            25  
Standard Expiration?       08/05/2018  
Oper No?                    666666

Flow Cal. Date:            08/25/2016  
Slope                      701  
Intercept                  -701324

IR Calibration Date:       08/25/2016  
                                    3um                      9um

-----  
0th Coef(\*100):        -8608                      -21568  
1st Coef(\*100):        258710                     128604  
2nd Coef(\*100):        2478                       1279  
H2O adj(mg/l\*10k):    381                        528

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:            PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = BSPD      8164.14.00 09/16  
04/20/2018      12:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:59
02 Std. Sol.	0.021	13:00
03 Room Air	0.000	13:01
04 Std. Sol.	0.022	13:01
05 Room Air	0.000	13:02
06 Std. Sol.	0.021	13:03
07 Room Air	0.000	13:03

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 201610A

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Low AC*

*0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = BSPD      8164.14.00 09/16  
04/20/2018      13:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:04
02 Std. Sol.	0.294	13:05
03 Room Air	0.000	13:05
04 Std. Sol.	0.294	13:06
05 Room Air	0.000	13:07
06 Std. Sol.	0.295	13:07
07 Room Air	0.000	13:08

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006675  
Location = BSPD                    8164.14.00 09/16  
04/20/2018                            13:09

Test	AC	Time
01 Room Air	0.000	13:10
02 *Subject Test	INT*	13:11
03 Room Air	0.000	13:11

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male                      Weight = 120  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821449  
Lot No = 17316080A2  
Cyl No = 25  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10 AC + 0.05% ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-006675  
Location = BSPD 8164.14.00 09/16  
04/20/2018 13:12

Test	AC	Time
01 Diagnostic	OK	13:13
02 Room Air	0.000	13:14
03 *Subject Test	RFI*	13:14
04 Room Air	0.000	13:14

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 03/01/1982  
Sub Sex = Male Weight = 120  
Test = DUI Cit = RFI CHECK  
Dr. Lic. = ND/DIS821449  
Lot No = 17316080A2  
Cyl No = 25  
Expiration Date = 08/05/2018  
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = BSPD      8164.14.00 09/16  
04/20/2018      13:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:16
02 Std. Gas	0.080	13:16
03 Room Air	0.000	13:17
04 Std. Gas	0.081	13:17
05 Room Air	0.000	13:17
06 Std. Gas	0.080	13:18
07 Room Air	0.000	13:18

Lot No = 17316080A2

Cyl No = 25

Exp Date = 08/05/2018

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = BSPD      8164.14.00 09/16  
04/20/2018      13:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Gas	0.081	13:19
03 Room Air	0.000	13:20
04 Std. Gas	0.081	13:20
05 Room Air	0.000	13:21
06 Std. Gas	0.081	13:21
07 Room Air	0.000	13:22

Lot No = 17316080A2

Cyl No = 25

Exp Date = 08/05/2018

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006675  
Location = BSPD      8164.14.00 09/16  
04/20/2018      13:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:22
02 Std. Gas	0.081	13:23
03 Room Air	0.000	13:23
04 Std. Gas	0.081	13:24
05 Room Air	0.000	13:24
06 Std. Gas	0.080	13:25
07 Room Air	0.000	13:25

Lot No = 17316080A2

Cyl No = 25

Exp Date = 08/05/2018

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000