



# ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

|   |   |                                   |
|---|---|-----------------------------------|
| Chemical Test Operator Name (Print)<br>OLSON, CURTIS E. |   |                                   |
| Location<br>GARRISON                                    | Intoxilyzer® Serial Number<br>80-006670 |                                   |
| Gas Lot Number<br>34917080A3                            | Gas Cylinder Number<br>11               | Gas Expiration Date<br>02-05-2020 |

Check When Done:



- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).

ACA test Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

was performed  
24 Apr 18 (Ryden)

A. 0. 081 AC  
B. 0. 081 AC  
C. 0. 081 AC

- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

|   |              |
|---|--------------|
| Chemical Test Operator Signature         | Date 4-20-18 |
| Reviewed By (Crime Laboratory Use Only)  | Date 4-24-18 |

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006670  
Location = GARR      8164.14.00 09/16  
04/20/2018      19:39

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 19:39 |
| 02 Std. Gas | 0.081 | 19:40 |
| 03 Room Air | 0.000 | 19:40 |
| 04 Std. Gas | 0.081 | 19:41 |
| 05 Room Air | 0.000 | 19:41 |
| 06 Std. Gas | 0.081 | 19:41 |
| 07 Room Air | 0.000 | 19:42 |

Lot No = 34917080A3  
Cyl No = 11  
Exp Date = 02/05/2020  
County = 28      Oper No. = 059008



Operator Signature  
CURTIS OLSON

Remarks: