



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>OLSON, CURTIS E.</i>		
Location <i>WASHBURN</i>	Intoxilyzer® Serial Number <i>80-006669</i>	
Gas Lot Number <i>34917080A3</i>	Gas Cylinder Number <i>10</i>	Gas Expiration Date <i>02-05-2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 081 AC
 - B. 0. 080 AC
 - C. 0. 081 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>04-23-2018</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>04-26-2018</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
04/23/2018 15:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:52
02 Std. Gas	0.081	15:52
03 Room Air	0.000	15:53
04 Std. Gas	0.080	15:53
05 Room Air	0.000	15:54
06 Std. Gas	0.081	15:54
07 Room Air	0.000	15:54

Lot No = 34917080A3

Cyl No = 10

Exp Date = 02/05/2020

County = 28

Oper No. = 059008



Operator Signature
CURTIS OLSON

Remarks:

Form 106-I8000