

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006676

Location: MAND

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 280 psi Regulator: 300 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.026 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 17316080 A2 Cyl #: 33 Exp. Date: 8/5/18

Test 1	<u>0.082</u>	Test 4	<u>0.082</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ehr
Breath Analyst Signature

4.27.2018
Date

Roberta Huggins
Reviewed by

4.27.2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:30
02 Std. Sol.	0.018	13:31
03 Room Air	0.000	13:32
04 Std. Sol.	0.018	13:32
05 Room Air	0.000	13:33
06 Std. Sol.	0.019	13:34
07 Room Air	0.000	13:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = ~~17350~~ 201610A CEE
County = 30 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:37
02 Std. Sol.	0.293	13:38
03 Room Air	0.000	13:39
04 Std. Sol.	0.296	13:39
05 Room Air	0.000	13:40
06 Std. Sol.	0.296	13:41
07 Room Air	0.000	13:41

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 30

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:22

Test	AC	Time
01 Room Air	0.000	13:23
02 *Subject Test	INT*	13:23
03 Room Air	0.000	13:24

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821449
Lot No = 17316080A2
Cyl No = 33
Expiration Date = 08/05/2018
County = 30 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AL + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:25

Test	AC	Time
01 Diagnostic	OK	13:26
02 Room Air	0.000	13:26
03 *Subject Test	RFI*	13:27
04 Room Air	0.000	13:27

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 03/01/1982
Sub Sex = Male Weight = 120
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821449
Lot No = 17316080A2
Cyl No = 33
Expiration Date = 08/05/2018
County = 30 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Gas	0.082	13:19
03 Room Air	0.000	13:20
04 Std. Gas	0.081	13:20
05 Room Air	0.000	13:20
06 Std. Gas	0.081	13:21
07 Room Air	0.000	13:21

Lot No = 17316080A2
Cyl No = 33
Exp Date = 08/05/2018
County = 30 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Gas	0.082	13:44
03 Room Air	0.000	13:45
04 Std. Gas	0.081	13:45
05 Room Air	0.000	13:46
06 Std. Gas	0.081	13:46
07 Room Air	0.000	13:47

Lot No = 17316080A2

Cyl No = 33

Exp Date = 08/05/2018

County = 30

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006676
Location = MAND 8164.14.00 09/16
04/27/2018 13:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Gas	0.081	13:48
03 Room Air	0.000	13:48
04 Std. Gas	0.081	13:49
05 Room Air	0.000	13:49
06 Std. Gas	0.081	13:50
07 Room Air	0.000	13:50

Lot No = 17316080A2
Cyl No = 33
Exp Date = 08/05/2018
County = ~~08~~ *30 CEE* Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000