



ETHANOL GAS STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Travis Collins</i>		
Location <i>HETT</i>	Intoxilyzer® Serial Number <i>80-006510</i>	
Gas Lot Number <i>34917080A3</i>	Gas Cylinder Number <i>14</i>	Gas Expiration Date <i>02/05/2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 0 8 0 AC
 - B. 0. 0 8 0 AC
 - C. 0. 0 8 1 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>T Collins</i>	Date <i>04/27/2018</i>
Reviewed By (Crime Laboratory Use Only) <i>Robert Sanchez - Nemo</i>	Date <i>05/08/2018</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58502

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN B0-006510
Location = HETT 8164.14.00 09/16
04/27/2018 18:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:38
02 Std. Gas	0.080	18:38
03 Room Air	0.000	18:39
04 Std. Gas	0.080	18:39
05 Room Air	0.000	18:40
06 Std. Gas	0.081	18:40
07 Room Air	0.000	18:41

Lot No = 34917080A3
Cyl No = 14
Exp Date = 02/05/2020
County = 01 *RSN* Oper No. = 020533

J Collins

Operator Signature
TRAVIS COLLINS

Remarks: *cyl. change*