

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006672

Location: NEW TOWN P.D.
NTPD

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 922 psi Regulator: 925 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10-4-18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS #5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 0341708043 Cyl #: 30 Exp. Date: 3/5/19

Test 1	<u>0.082</u>	Test 4	<u>0.081</u>	Test 7	<u>0.080</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.081</u>
Average <u>0.081</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Edr
Breath Analyst Signature

5-7-2018
Date

Roberta Henrique-Munim
Reviewed by

14 May 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 30
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 07/05/2017
Slope 705
Intercept -687158

IR Calibration Date: 07/05/2017

	3um	9um
0th Coef(*100):	-22113	-17955
1st Coef(*100):	271833	130091
2nd Coef(*100):	2129	1363
H2O adj(mg/l*10k):	691	568

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.018	13:52
03 Room Air	0.000	13:53
04 Std. Sol.	0.018	13:53
05 Room Air	0.000	13:54
06 Std. Sol.	0.018	13:55
07 Room Air	0.000	13:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3377

Std Sol No = 210610A

County = 31

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Sol.	0.300	13:57
03 Room Air	0.000	13:57
04 Std. Sol.	0.297	13:58
05 Room Air	0.000	13:59
06 Std. Sol.	0.299	13:59
07 Room Air	0.000	14:00

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 17350
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:45

Test	AC	Time
01 Room Air	0.000	13:46
02 *Subject Test	INT*	13:46
03 Room Air	0.000	13:47

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 30
Expiration Date = 03/05/2019
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:36

Test	AC	Time
01 Diagnostic	OK	13:37
02 Room Air	0.000	13:37
03 *Subject Test	RFI*	13:37
04 Room Air	0.000	13:38

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 30
Expiration Date = 03/05/2019
County = 31 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Gas	0.082	13:42
03 Room Air	0.000	13:43
04 Std. Gas	0.082	13:43
05 Room Air	0.000	13:44
06 Std. Gas	0.081	13:44
07 Room Air	0.000	13:45

Lot No = 03417080A3
Cyl No = 30
Exp Date = 03/05/2019
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 13:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Gas	0.081	13:48
03 Room Air	0.000	13:48
04 Std. Gas	0.081	13:49
05 Room Air	0.000	13:49
06 Std. Gas	0.081	13:50
07 Room Air	0.000	13:50

Lot No = 03417080A3
Cyl No = 30
Exp Date = 03/05/2019
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006672
Location = NTPD 8164.14.00 09/16
05/07/2018 14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Gas	0.080	14:02
03 Room Air	0.000	14:02
04 Std. Gas	0.081	14:02
05 Room Air	0.000	14:03
06 Std. Gas	0.081	14:03
07 Room Air	0.000	14:04

Lot No = 03417080A3
Cyl No = 30
Exp Date = 03/05/2019
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000