

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006669

Location: WASH
MCLEAN COUNTY
S.O.

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 997 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3377 Lot #: 201610A Exp. Date: 10.4.18
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: DR5191 Lot #: 17350 Exp. Date: 10.11.19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS#5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34917080A3 Cyl #: 10 Exp. Date: 2/5/20

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.080</u>
Test 2	<u>0.080</u>	Test 5	<u>0.079</u>	Test 8	<u>0.086</u>
Test 3	<u>0.080</u>	Test 6	<u>0.079</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ehr
Breath Analyst Signature

5-4-2018
Date

Roberta Muecke-Timmer
Reviewed by

5-9-2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34917080A3
Standard Cyl #? 10
Standard Expiration? 02/05/2020
Oper No? 666666

Flow Cal. Date: 08/15/2016
Slope 678
Intercept -615536

IR Calibration Date: 08/15/2016
 3um 9um

0th Coef(*100): -19192 -25323
1st Coef(*100): 259438 128900
2nd Coef(*100): 2747 1289
H2O adj(mg/l*10k): 613 563

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Sol.	0.017	14:27
03 Room Air	0.000	14:28
04 Std. Sol.	0.017	14:28
05 Room Air	0.000	14:29
06 Std. Sol.	0.018	14:30
07 Room Air	0.000	14:30

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 201610A
County = 28 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:21
02 Std. Sol.	0.292	14:21
03 Room Air	0.000	14:22
04 Std. Sol.	0.292	14:23
05 Room Air	0.000	14:23
06 Std. Sol.	0.292	14:24
07 Room Air	0.000	14:25

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 17350

County = 28

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC

0.300 ~~0.020~~ AC
CEE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:05

Test	AC	Time
01 Room Air	0.000	14:06
02 *Subject Test	INT*	14:07
03 Room Air	0.000	14:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 10
Expiration Date = 02/05/2020
County = 28 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:03

Test	AC	Time
01 Diagnostic	OK	14:04
02 Room Air	0.000	14:05
03 *Subject Test	RFI*	14:05
04 Room Air	0.000	14:05

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 10
Expiration Date = 02/05/2020
County = 28 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:08
02 Std. Gas	0.079	14:09
03 Room Air	0.000	14:09
04 Std. Gas	0.080	14:09
05 Room Air	0.000	14:10
06 Std. Gas	0.080	14:10
07 Room Air	0.000	14:11

Lot No = 34917080A3
Cyl No = 10
Exp Date = 02/05/2020
County = 28 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Gas	0.080	14:12
03 Room Air	0.000	14:12
04 Std. Gas	0.079	14:13
05 Room Air	0.000	14:13
06 Std. Gas	0.079	14:14
07 Room Air	0.000	14:14

Lot No = 34917080A3
Cyl No = 10
Exp Date = 02/05/2020
County = 28 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006669
Location = WASH 8164.14.00 09/16
05/04/2018 14:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:31
02 Std. Gas	0.080	14:32
03 Room Air	0.000	14:32
04 Std. Gas	0.080	14:33
05 Room Air	0.000	14:33
06 Std. Gas	0.080	14:33
07 Room Air	0.000	14:34

Lot No = 34917080A3
Cyl No = 10
Exp Date = 02/05/2020
County = 28 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC